

SOLUTION CHANGE

DataMaster DMT: 100156
Location: FAIR HAVEN PD
Date: 05/18/2020
Performed by: DALE H KERBER



Accuracy and Precision Check

Concentration = 0.101 g/210L
Lot # = 19-16-100
Range = 0.096 - 0.106
Average = 0.097 g/210L
Std Dev = 0.0004

Handwritten: JSD 5/18/2020

Simulator Temperature: 34.0°C

Performed by

Handwritten: Dale H Kerber

Date

05/18/2020