



DMT Technical Support Inquiry

DMT Serial #: 100156
Agency/Site: Fair Haven PD
Date: 5/29/2021
DMT Supervisor: Dale Kerber

Problem:

Multiple simulator time out errors were obtained trying to perform the RPC. JSD spoke on the phone with DMT supervisor to troubleshoot. The simulator temperature was stable and the paddle was spinning properly. DMT Supervisor had previously removed simulator head and reconnected to no avail. The large o-ring was also checked and the quick connects were confirmed to be secure. The detector voltage at filter 1 was fluctuating 4-6 mV with pump off, pump on, and when the sim valve option was toggled. Determined to be an instrument issue. DMT was returned to VFL for service.

Work Performed:

Prior to the DMT's removal from the agency on 6/2/2021, the voltage plot indicated that the voltage was indeed fluctuating. The seal on the simulator was also tested to verify that the jar was securely fastened. On 6/3/2021, the detector was replaced. The detector block was taken apart and cleaned with compressed air before being put back together. A calibration was run for testing purposes, followed by A&P testing to ensure that the simulator was no longer timing out. 52 filter wheel cycles were run without error. The voltages were optimized and allowed to stabilize. Voltage differed by >3 mV over the course of a 1 minute purge. Adjusted wiring with no change. Replaced the controller board and allowed voltages to stabilize. The detector voltage was optimized and a successful calibration adjustment and certification were performed. Instrument is now ready for installation at agency.

Performed On Site ☒

Performed In house ☒

Performed By: *SS* Sarah Stratton and *JD* Jeff Dukette

Date: 6/9/2021

Technical Reviewer: *AS*

Date: 6/11/21

Administrative Reviewer: *AS*

Date: 6/11/21

Director Reviewer: *UCA*

Date: 6/11/2021

CALIBRATION REPORT

DataMaster DMT: 100156
Calibration Date: 06/09/2021
Calibrated by: SARAH STRATTON
Lot: 21-02-100



Ca	=	0.100		
CAL	=	0.953220	0.800	<= CAL < 1.200
b1	=	0.000	0.000	<= b1 < 0.004
b2	=	0.005	0.002	<= b2 < 0.010
b3	=	0.000	0.000	<= b3 < 0.004
Xq	=	0.108	0.050	<= Xq < 0.200
a21	=	1.192483	1.050	<= a21 < 1.300
a31	=	0.455670	0.300	<= a31 < 0.800

DI Water


BL-DMT01072020 34.00°C SJS 6/9/2021

Performed by

SJS SJS

Date

06/09/2021



ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMaster DMT: 100156

Date: 06/09/2021

Time: 09:20:44

SUPERVISOR NAME:


SGS CAL CHECK

SOLUTION LOT #: ACS202010E

SOLUTION CONCENTRATION: 0.100

BLANK TEST	0.000	09:21
CALIBRATION CHECK	PASSED	09:21
SIMULATOR VAPOR 34.0°C	0.098	09:22
SIMULATOR VAPOR 34.0°C	0.098	09:23
SIMULATOR VAPOR 34.0°C	0.098	09:24
SIMULATOR VAPOR 34.0°C	0.098	09:25
SIMULATOR VAPOR 34.0°C	0.098	09:26
SIMULATOR VAPOR 34.0°C	0.098	09:27
SIMULATOR VAPOR 34.0°C	0.098	09:28
SIMULATOR VAPOR 34.0°C	0.098	09:29
SIMULATOR VAPOR 34.0°C	0.098	09:31
SIMULATOR VAPOR 34.0°C	0.098	09:32
BLANK TEST	0.000	09:33

Average = 0.098
Std Dev = 0.0000



CERTIFICATION REPORT

DataMaster DMT: 100156
Calibration Date: 06/09/2021
Certification Date: 06/09/2021
Certified by: SARAH STRATTON



Diagnostic Results

VERSIONS

DMT: 2.06
PIC: 3.07
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.7°C
Breath Tube = 42.8°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.62 V
Cooler Voltage = 1.66 V
Bias Voltage = 80 V
Chopper Freq = 522 Hz

PUMP INFO

Flow Rate = 5.622 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX (V)	0.0244	0.0287
MIN (V)	0.0228	0.0268

FILTER INFO

Filter 1	0.028	Zero = true
Filter 2	0.318	Zero = true
Filter 3	0.990	Zero = true

CALIBRATION CHECK

Xq = 0.108 0.08%

Options

OPTIONS

Units

Alcohol..... g/210L

Simulator

Tolerance Check..... yes
Standard Type..... wet
Nominal..... 0.100
Digital Simulator..... Guth

Subject

Ask Questions..... yes
Number of Tests..... 2
Alcohol Display..... yes
Volume Display..... yes
Query Refusal..... yes
Copies..... 3
Simulator Before..... yes
Simulator Between..... yes
Simulator After..... no
Observation Time..... 0

Supervisor

Number of Tests..... 10

Calibration

Standard Type..... wet
Nominal..... 0.100
Nominal (Dry Gas)..... 0.100
Number of Tests..... 1


Printer

Printer On..... yes
Software Configuration
Data Collection..... yes

Linearity Check Results

True Value	Reported Average	Std Dev
0.020 g/210L	0.019 g/210L	0.0000
Lot # 21-01-020		
0.080 g/210L	0.077 g/210L	0.0005
Lot # 20-12-080		
0.162 g/210L	0.157 g/210L	0.0004
Lot # 21-07-160		
0.371 g/210L	0.354 g/210L	0.0000
Lot # 20-16-360		

$R^2 = 0.9999$



Acetone Interference Test

Lot # 20-13-08A2


Interference Detected

Mouth Alcohol Test

Mouth Alcohol Detected


RF Detection Test

Passed



Sample Acceptance Test

Passed



CERTIFICATION PASSED

Performed by

SJS SJS

Date

06/09/2021

Reviewed by

[Signature]

Date

6/11/21

DMT Serial Number #100156

Page 2 of 2

06/09/2021 12:32 PM

100156

SJS

5
3
SJS 6/10/2021

DIAGNOSTIC RESULT



DataMaster DMT: 100156
Location:
Calibration Date: 06/03/2021
Certification Date:
Installation Date:
Test Date: 06/09/2021
Test Time: 09:03:34

VERSIONS

DMT: 2.06
PIC: 3.07
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.9°C
Breath Tube = 46.1°C
Digital Sim = 0.0°C * Sim not plugged in for diags SJS 6/9/2021

SETTINGS

Lamp Voltage = 1.62 V
Cooler Voltage = 1.66 V
Bias Voltage = 80 V
Chopper Freq = 523 Hz

PUMP INFO

Flow Rate = 5.506 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX (V)	-0.0028	0.0022
MIN (V)	-0.0041	0.0008


FILTER INFO

Filter 1	0.001	Zero = true
Filter 2	0.293	Zero = true
Filter 3	0.957	Zero = true

CALIBRATION CHECK

Xq = 0.107 0.66%

Diag run after changing controller board. SJS 6/10/2021



ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMaster DMT: 100156

Date: 06/03/2021

Time: 15:39:10

SUPERVISOR NAME:


SARAH STRATTON

SOLUTION LOT #: TEST

SOLUTION CONCENTRATION: 0.100

BLANK TEST	0.000	15:40
CALIBRATION CHECK	PASSED	15:40
SIMULATOR VAPOR 34.0°C	0.100	15:40
SIMULATOR VAPOR 34.0°C	0.099	15:41
SIMULATOR VAPOR 34.0°C	0.099	15:42
SIMULATOR VAPOR 34.0°C	0.099	15:43
SIMULATOR VAPOR 34.0°C	0.099	15:44
SIMULATOR VAPOR 34.0°C	0.099	15:45
SIMULATOR VAPOR 34.0°C	0.099	15:46
SIMULATOR VAPOR 34.0°C	0.099	15:47
SIMULATOR VAPOR 34.0°C	0.099	15:48
SIMULATOR VAPOR 34.0°C	0.099	15:49
BLANK TEST	0.000	15:50

Average = 0.099
Std Dev = 0.0003



CALIBRATION REPORT

DataMaster DMT: 100156
Calibration Date: 06/03/2021
Calibrated by: SARAH STRATTON
Lot: TEST



Ca	=	0.100		
CAL	=	0.953083	0.800	<= CAL < 1.200
b1	=	0.000	0.000	<= b1 < 0.004
b2	=	0.004	0.002	<= b2 < 0.010
b3	=	0.000	0.000	<= b3 < 0.004
Xq	=	0.107	0.050	<= Xq < 0.200
a21	=	1.188700	1.050	<= a21 < 1.300
a31	=	0.458702	0.300	<= a31 < 0.800

Calibration for testing purposes only. SJS 6/3/2021

Performed by

SJS

Date

06/03/2021

DIAGNOSTIC RESULT



DataMaster DMT: 100156
Location: FAIR HAVEN PD
Calibration Date: 03/10/2017
Certification Date: 03/10/2017
Installation Date: 03/13/2017
Test Date: 06/03/2021
Test Time: 15:31:47

VERSIONS

DMT: 2.06
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.7°C
Breath Tube = 45.8°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.61 V
Cooler Voltage = 1.65 V
Bias Voltage = 80 V
Chopper Freq = 531 Hz

PUMP INFO

Flow Rate = 5.484 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX (V)	-0.0376	-0.0333
MIN (V)	-0.0402	-0.0348

FILTER INFO

Filter 1	-0.035	Zero = true
Filter 2	0.261	Zero = true
Filter 3	0.912	Zero = true

CALIBRATION CHECK

Xq = 0.106 8.05%

Diag run after blowing compressed air on quartz disc. SJS 6/3/2021

DIAGNOSTIC RESULT

DataMaster DMT: 100156
Location: FAIR HAVEN PD
Calibration Date: 03/10/2017
Certification Date: 03/10/2017
Installation Date: 03/13/2017
Test Date: 06/03/2021
Test Time: 13:18:10



VERSIONS

DMT: 2.06
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.6°C
Breath Tube = 44.4°C
Digital Sim = 0.0°C

SETTINGS

Lamp Voltage = 1.61 V
Cooler Voltage = 1.66 V
Bias Voltage = 80 V
Chopper Freq = 529 Hz

PUMP INFO

Flow Rate = 5.551 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX (V)	-0.1230	-0.1171
MIN (V)	-0.1244	-0.1183

FILTER INFO

Filter 1	-0.117	Zero = true
Filter 2	0.188	Zero = true
Filter 3	0.806	Zero = true

CALIBRATION CHECK

Xq = 0.105 6.70%

Diag run after replacing the detector. SJS 6/3/2021

ROUTINE PERFORMANCE CHECK REPORT



DataMaster DMT: 100156
Location: FAIR HAVEN PD
Calibration Date: 03/10/2017
Certification Date: 03/10/2017
Installation Date: 03/13/2017
RPC Date: 05/30/2021
Supervisor Name: DALE H KERBER

Diagnostic Results

VERSIONS

DMT: 2.06
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.9°C
Breath Tube = 45.7°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.64 V
Cooler Voltage = 1.71 V
Bias Voltage = 80 V
Chopper Freq = 538 Hz

PUMP INFO

Flow Rate = 5.526 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX(V)	0.3937	0.3970
MIN(V)	0.3890	0.3909

FILTER INFO

Filter 1	0.397	Zero = true
Filter 2	0.659	Zero = true
Filter 3	1.294	Zero = true

CALIBRATION CHECK

Xq = 0.096 1.34%

Routine Performance Check Failed

Reason: ACCURACY AND PRECISION CHECK FAILED

Accuracy and Precision Check

Concentration

Lot #

Range

Average

Std Dev

Simulator Timed Out

Performed by

Dale Kerber

Date

05/30/2021

Reviewed by

Date

ROUTINE PERFORMANCE CHECK REPORT



DataMaster DMT: 100156
Location: FAIR HAVEN PD
Calibration Date: 03/10/2017
Certification Date: 03/10/2017
Installation Date: 03/13/2017
RPC Date: 05/29/2021
Supervisor Name: DALE H KERBER

Diagnostic Results

VERSIONS

DMT: 2.06
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.9°C
Breath Tube = 46.8°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.64 V
Cooler Voltage = 1.71 V
Bias Voltage = 80 V
Chopper Freq = 538 Hz

PUMP INFO

Flow Rate = 5.504 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX(V)	0.4098	0.4136
MIN(V)	0.4051	0.4089

FILTER INFO

Filter 1	0.413	Zero = true
Filter 2	0.671	Zero = true
Filter 3	1.309	Zero = true

CALIBRATION CHECK

Xq = 0.096 1.34%

Routine Performance Check Failed

Reason: ACCURACY AND PRECISION CHECK FAILED

Accuracy and Precision Check

Concentration

Lot #

Range

Average

Std Dev

Simulator Timed Out

Performed by

Dale Kerber

Date

05/29/2021

Reviewed by

Date

ROUTINE PERFORMANCE CHECK REPORT



DataMaster DMT: 100156
Location: FAIR HAVEN PD
Calibration Date: 03/10/2017
Certification Date: 03/10/2017
Installation Date: 03/13/2017
RPC Date: 05/29/2021
Supervisor Name: DALE H KERBER

Diagnostic Results

VERSIONS

DMT: 2.06
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.9°C
Breath Tube = 46.8°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.64 V
Cooler Voltage = 1.71 V
Bias Voltage = 80 V
Chopper Freq = 538 Hz

PUMP INFO

Flow Rate = 5.519 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX(V)	0.4069	0.4100
MIN(V)	0.4022	0.4043

FILTER INFO

Filter 1	0.409	Zero = true
Filter 2	0.670	Zero = true
Filter 3	1.307	Zero = true

CALIBRATION CHECK

Xq = 0.098 0.71%

Routine Performance Check Failed

Reason: ACCURACY AND PRECISION CHECK FAILED

Accuracy and Precision Check

Concentration

Lot #

Range

Average

Std Dev

Simulator Timed Out

Performed by

Dale Kerber

Date

05/29/2021

Reviewed by

Date

ROUTINE PERFORMANCE CHECK REPORT



DataMaster DMT: 100156
Location: FAIR HAVEN PD
Calibration Date: 03/10/2017
Certification Date: 03/10/2017
Installation Date: 03/13/2017
RPC Date: 05/29/2021
Supervisor Name: DALE H KERBER

Diagnostic Results

VERSIONS

DMT: 2.06
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.9°C
Breath Tube = 47.2°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.64 V
Cooler Voltage = 1.71 V
Bias Voltage = 80 V
Chopper Freq = 538 Hz

PUMP INFO

Flow Rate = 5.486 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX(V)	0.4020	0.4061
MIN(V)	0.3976	0.4014

FILTER INFO

Filter 1	0.402	Zero = true
Filter 2	0.667	Zero = true
Filter 3	1.307	Zero = true

CALIBRATION CHECK

Xq = 0.097 0.54%

Routine Performance Check Failed

Reason: ACCURACY AND PRECISION CHECK FAILED

Accuracy and Precision Check

Concentration

Lot #

Range

Average

Std Dev

Simulator Timed Out

Performed by

Dale Kerber

Date

05/29/2021

Reviewed by

Date

ROUTINE PERFORMANCE CHECK REPORT

DataMaster DMT: 100156
Location: FAIR HAVEN PD
Calibration Date: 03/10/2017
Certification Date: 03/10/2017
Installation Date: 03/13/2017
RPC Date: 05/29/2021
Supervisor Name: DALE H KERBER



Diagnostic Results

VERSIONS

DMT: 2.06
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.7°C
Breath Tube = 47.8°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.64 V
Cooler Voltage = 1.71 V
Bias Voltage = 80 V
Chopper Freq = 537 Hz

PUMP INFO

Flow Rate = 5.570 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX(V)	0.3978	0.4014
MIN(V)	0.3937	0.3968

FILTER INFO

Filter 1	0.399	Zero = true
Filter 2	0.661	Zero = true
Filter 3	1.301	Zero = true

CALIBRATION CHECK

Xq = 0.098 0.09%

Routine Performance Check Failed

Reason: ACCURACY AND PRECISION CHECK FAILED

Accuracy and Precision Check

Concentration

Lot #

Range

Average

Std Dev

Simulator Timed Out

Performed by

Dale Kerber

Date

05/29/2021

Reviewed by

Date

ROUTINE PERFORMANCE CHECK REPORT



DataMaster DMT: 100156
Location: FAIR HAVEN PD
Calibration Date: 03/10/2017
Certification Date: 03/10/2017
Installation Date: 03/13/2017
RPC Date: 05/31/2021
Supervisor Name: DALE H KERBER

Diagnostic Results

VERSIONS

DMT: 2.06
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.9°C
Breath Tube = 46.0°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.64 V
Cooler Voltage = 1.71 V
Bias Voltage = 80 V
Chopper Freq = 538 Hz

PUMP INFO

Flow Rate = 5.490 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX(V)	0.3984	0.4014
MIN(V)	0.3923	0.3959

FILTER INFO

Filter 1	0.398	Zero = true
Filter 2	0.661	Zero = true
Filter 3	1.296	Zero = true

CALIBRATION CHECK

Xq = 0.097 0.62%

Routine Performance Check Failed

Reason: ACCURACY AND PRECISION CHECK FAILED

Accuracy and Precision Check

Concentration

Lot #

Range

Average

Std Dev

Simulator Timed Out

Performed by

Dale Kerber

Date

05/31/2021

Reviewed by

Date

ROUTINE PERFORMANCE CHECK REPORT

DataMaster DMT: 100156
Location: FAIR HAVEN PD
Calibration Date: 03/10/2017
Certification Date: 03/10/2017
Installation Date: 03/13/2017
RPC Date: 05/29/2021
Supervisor Name: DALE H KERBER



Diagnostic Results

VERSIONS

DMT: 2.06
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.8°C
Breath Tube = 47.9°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.64 V
Cooler Voltage = 1.71 V
Bias Voltage = 80 V
Chopper Freq = 537 Hz

PUMP INFO

Flow Rate = 5.440 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX(V)	0.3994	0.4006
MIN(V)	0.3923	0.3957

FILTER INFO

Filter 1	0.397	Zero = true
Filter 2	0.661	Zero = true
Filter 3	1.299	Zero = true

CALIBRATION CHECK

Xq = 0.098 0.00%

Routine Performance Check Failed

Reason: ACCURACY AND PRECISION CHECK FAILED

Accuracy and Precision Check

Concentration

Lot #

Range

Average

Std Dev

Simulator Timed Out

Performed by

Dale Kerber

Date

05/29/2021

Reviewed by

Date

Call Log

Person	Date	Type	VFL Staff	Case/DMT	Content
Dale Kerber	6/1/2021	Phone	JSD	Fair Haven PD	<p>They were receiving simulator time out errors. Spoke on the phone with Dale to troubleshoot</p> <p>Is simulator temperature stable - stable</p> <p>Is the paddle spinning properly - paddle is spinning</p> <p>Remove simulator and re-screw on - done multiple times</p> <p>Remove simulator and check o ring - he removed and put back</p> <p>Check connection pieces to see if they are loose - All good</p> <p>Detector voltage without pump on - 0.399 - 0.404</p> <p>Detector voltage with pump on - 0.390-0.394</p> <p>Detector voltage with sim valve toggled - 0.616 - 0.622</p>

100156

8298

Re: Fair Haven Datamaster

Kerber, Dale <Dale.Kerber@vermont.gov>

Tue 2021-06-01 10:21 AM

To: DPS - DMT <DPS.DMT@vermont.gov>

OK I'll head to the office right now I'll give you a call

Get [Outlook for iOS](#)

From: DPS - DMT <DPS.DMT@vermont.gov>

Sent: Tuesday, June 1, 2021 10:20:51 AM

To: Kerber, Dale <Dale.Kerber@vermont.gov>

Subject: Re: Fair Haven Datamaster

If you don't mind, that would be great.

You can reach me directly at 802-585-5843

Thanks!

Jeff

From: Kerber, Dale <Dale.Kerber@vermont.gov>

Sent: Tuesday, June 1, 2021 10:19 AM

To: DPS - DMT <DPS.DMT@vermont.gov>

Subject: Re: Fair Haven Datamaster

I'm not working in the office today but I can be there in 10-15 minutes and take a call, Dale

Get [Outlook for iOS](#)

From: DPS - DMT <DPS.DMT@vermont.gov>

Sent: Tuesday, June 1, 2021 9:34:36 AM

To: Kerber, Dale <Dale.Kerber@vermont.gov>

Subject: Re: Fair Haven Datamaster

Hey Dale,

This happens periodically especially after a solution change, but it shouldn't impact the results like that. I pulled the raw data this morning and observed that fluctuation in the alcohol line we discussed. This is either due to leaky simulator or some voltage instability. Do you have any time today to take a phone call while at the DMT so I can run you through some quick troubleshooting steps to help us diagnose the issue? We will be coming down to that area this week to check out a different instrument and it would be helpful to know if it's something we can possibly fix in the field vs bringing it back to the lab.

Thanks!

Jeff

From: Kerber, Dale <Dale.Kerber@vermont.gov>
Sent: Sunday, May 30, 2021 5:35 PM
To: DPS - DMT <DPS.DMT@vermont.gov>
Subject: Re: Fair Haven Datamaster

I gave it another try today and same results. I did notice some drops of solution on the nipple that sticks out of the dmt when you disconnect the simulator solution. I've never noticed this before. Is that normal? Dale K

Get [Outlook for iOS](#)

From: Kerber, Dale <Dale.Kerber@vermont.gov>
Sent: Saturday, May 29, 2021 3:19:32 PM
To: DPS - DMT <DPS.DMT@vermont.gov>
Subject: Re: Fair Haven Datamaster

I have done that and yes the line goes right up to .10 and fluctuates right there until it times out. I'm going to give it another try later, Dale Kerber

Get [Outlook for iOS](#)

From: DPS - DMT <DPS.DMT@vermont.gov>
Sent: Saturday, May 29, 2021 3:04:45 PM
To: Kerber, Dale <Dale.Kerber@vermont.gov>
Subject: Re: Fair Haven Datamaster

Hey Dale,

Have you tried taking off the top of the simulator and re threading it? Typically, a sim time out error results from a leaky simulator. Is the black line approaching 0.1 on the screen but just not settling or is it far from the 0.1 concentration? If it's close to 0.1, Is it fluctuating significantly? Like it looks similar to a sine wave.

Get [Outlook for iOS](#)

From: Kerber, Dale <Dale.Kerber@vermont.gov>
Sent: Saturday, May 29, 2021 1:13:33 PM
To: DPS - DMT <DPS.DMT@vermont.gov>
Subject: Fair Haven Datamaster

I attempted to do the RPC several times this morning and afternoon and each time The Simulator Solution test timed out each time. I have attempted all the remedies listed in the manual to no avail.

I have taken the DMT out of service until I hear from you.

Thanks,

Dale Kerber

6/8/2021

Mail - DPS - DMT - Outlook

Sergeant, Fair Haven Police Department
3 North Park Place,
Fair Haven, VT 05743
PH: (802)265-7839
Cell: (802)558-2518