

ANNUAL PREVENTATIVE MAINTENANCE

DataMaster DMT: 100156
Location: FAIR HAVEN PD
Calibration Date: 06/09/2021
Certification Date: 06/09/2021
Installation Date: 06/11/2021
APM Date: 05/31/2023
Performed by: SARAH STRATTON



Diagnostic Results

VERSIONS

DMT: 2.06
PIC: 3.07
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.8°C
Breath Tube = 47.8°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.63 V
Cooler Voltage = 1.66 V
Bias Voltage = 80 V
Chopper Freq = 532 Hz

PUMP INFO

Flow Rate = 5.552 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX(V)	0.4128	0.4167
MIN(V)	0.4120	0.4157

FILTER INFO

Filter	Value	Zero
Filter 1	0.416	true
Filter 2	0.686	true
Filter 3	1.307	true

CALIBRATION CHECK

Xq = 0.107 0.08%

APM PASSED

Simulator Vapor Test


Concentration = 0.098 g/210L
Lot # = 22-14-100
Range = 0.093 - 0.103
Simulator Vapor = 0.098 g/210L

RF Detection Test

Passed
RF detected

Sample Acceptance Test

Passed



Performed by

SJS

Date

05/31/2023

Reviewed by

AS

Date

06/02/23

DMT Serial Number #100156

Page 1 of 1

100156

05/31/2023 1:22 PM

SJS

Page 1 of 2

VERMONT FORENSIC LABORATORY

Simulator Temperature Check

Simulator Serial Number:

MP3474

DMT Serial Number / Simulator Location:

100156 / Fair Haven PD

NIST Traceable Thermometer Serial Number:

21FMM02249

As Found

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator	33.85	34.01	0.16	N

* $\pm 0.10^{\circ}\text{C}$ agreement between thermometer and simulator

* $34.0 \pm 0.2^{\circ}\text{C}$ acceptable range

Was temperature adjustment performed?

(Y) / N

As Left (if applicable)

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator	34.00	34.00	0	Y

Notes:

Check Performed By:

SJS

Date:

5/31/2023

Reviewed by:

AS

Date:

6/6/23