

DATAMASTER DMT MAINTENANCE LOG

Colchester PD

Department/Agency

100157

DataMaster DMT Serial Number

2/17/12 - 350 1/28/25

2/17/12 - 1/15/25

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/17/12	DataMaster Simulator Solution Lot # 12-58-100 EtOH_{aq} Conc: 0.100 [Place Label Here] Range: 0.095 - 0.105 Preparation Date: 02/09/2012 Expires: 02/09/2013 OPENED: 2/17/12 Installation #1 Failed due To Blank error Installation #2 Passed. 2/17/12	Steven Harnois S. Bob
3/8/12	DataMaster Simulator Solution Lot # 12-58-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 02/09/2012 Expires: 02/09/2013 [Place Label Here] OPENED: 3/8/12 SW Moved Instrument To New Temporary Colchester Pol Location	Steven Harnois
4/3/12	DataMaster Simulator Solution Lot # 12-58-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 02/09/2012 Expires: 02/09/2013 OPENED: 4/3/2012 [Signature] Solution Change 0.095 0.099 4/3/12	[Signature] M193
5-30-12	DataMaster Simulator Solution Lot # 12-58-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 02/09/2012 Expires: 02/09/2013 OPENED: 5-30-12 Jones Solution Change / RPC #1 Failed Due To Solution Jar not Being Plugged Back in #2 Passed	[Signature] M181

DATAMASTER DMT MAINTENANCE LOG

Colchester PD

Department/Agency

100157

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/7/12	Software update and [For RPC or Simulator Solution Change Place Label Here] Lock charge	AB
8-13-12	Diagnostic Check [For RPC or Simulator Solution Change Place Label Here] Accuracy & Precision Test	AB
8/17/12	Solution Change [For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 12-70-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/17/2012 Expires: 04/17/2013 OPENED: 8-17-12 Jones	AB
9/12/12	DataMaster Simulator Solution [For RPC or Simulator Solution Change Place Label Here] Lot # 12-80-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/5/2012 Expires: 7/5/2013 OPENED: 9/12/12 RCD Install	R Driscoll

DATAMASTER DMT MAINTENANCE LOG

Colchester PD
Department/Agency

100157
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
01 Nov 12	DataMaster Simulator Solution Lot # 12-80-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/5/2012 Expires: 7/5/2013 OPENED: [For RPC Label Here] RPC	
2-3-13	DataMaster Simulator Solution Lot # 12-80-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/5/2012 Expires: 7/5/2013 OPENED: 2-3-13 CQ [For RPC or Label Here] RPC	
5-27-13	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 12-87-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/31/2012 Expires: 10/31/2013 OPENED: 5-27-13 M151	
8 July 13	RPC [For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 12-89-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 11/21/2012 Expires: 11/21/2013 OPENED: 8 July 13 168	

DATAMASTER DMT MAINTENANCE LOG

Colchester PD
Department/Agency

100157
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/11/13	[For RPC or Simulator Solution Change Place Label Here] Performed APM protocol	R. Driscoll DMT TECH
10/5/13	DataMaster Simulator Solution Lot # 13-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 05/10/2013 Expires: 05/10/2014 OPENED: 10-5-13 181 [For RP ce Label Here] RPC - RPC 1st Attempt Fail Due To not keying Portable and mistakenly pressing cancel @ RFI Test	C Jones MSI
1/20/14	DataMaster Simulator Solution Lot # 13-19-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/25/2013 Expires: 10/25/2014 OPENED: 1/20/14 193 [For RFI ace Label Here] RPC - Passed.	J. Lehmann DMT Supervisor
5-26-14	DataMaster Simulator Solution Lot # 13-21-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 12/20/2013 Expires: 12/20/2014 OPENED: 5-26-14 MSI Jones [For RPC e Label Here] Solution Change	C Jones MSI

DATAMASTER DMT MAINTENANCE LOG

Colchester Police Dept

Department/Agency

100157

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	[For RPC or Simulator Solution Change Place Label Here] <i>RPC Passed</i>	<i>Chris</i> <i>Lowe</i> <i>musl</i> <i>Patrol</i>
<i>7/11/14</i>	<i>Completed ARM Protocol</i> [For RPC or Simulator Solution Change Place Label Here] <i>Adjusted Clock ahead 13 mins to match correct time</i>	<i>R. Driscoll</i> <i>DMT TECH</i>
<i>7-27-14</i>	DataMaster Simulator Solution Lot # 13-21-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 12/20/2013 Expires: 12/20/2014 OPENED: <i>7-27-14 CMS</i> [For RPC or Simulator Solution Change Place Label Here] <i>Solution Change</i>	
	DataMaster Simulator Solution Lot # 13-21-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 12/20/2013 Expires: 12/20/2014 OPENED: <i>11-5-14 CMS</i> [For RPC or Simulator Solution Change Place Label Here] <i>RPC Fail solution out of Range. Placed out of service</i>	<i>Chris</i>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/6/14	DataMaster Simulator Solution Lot # 14-14-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/16/2014 Expires: 10/16/2015 [For RF] OPENED: 11-8-14 CMJ ace Label Here] RPC Pass Return To Service	C Jones MIS
12/23/14	DataMaster Simulator Solution Lot # 14-14-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/16/2014 Expires: 10/16/2015 [For RP] OPENED: 12/23/14 ROS ice Label Here] Broken simulator paddle. Replaced paddle and performed simulator solution change ROS 12/23/14	R. Dinscott DMT TECH.
01/10/15	DataMaster Simulator Solution Lot # 14-10-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 7/14/2014 Expires: 7/14/2015 [For RPC] OPENED: 11/10/15 SBL ice Label Here] Solution Change	J. Lehmann DMT Supervisor
2/20/15	DataMaster Simulator Solution Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 [For RP] OPENED: 2/20/15 SBL ice Label Here] RPC Passed.	J. Lehmann DMT Supervisor

DATAMASTER DMT MAINTENANCE LOG

Colchester PD
Department/Agency

100157
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/24/15	Alcohol "blip" during sample acceptance test on RPC [For RPC or Simulator Solution Change Place Label Here] DMT Supervisor tried running "check-in" but still getting "blip" (less than .005 g/210L). Checked 5-way valve operation and performed "check-in" test with all 4 breath samples with no "blips" RCD 2/24/15	R. Driscoll DMT TECH.
3/5/15	<u>DataMaster Simulator Solution</u> Lot # 15-01-100 EtOHaq Conc: 0.098 g/210L Range: 0.093 - 0.103 g/210L Preparation Date: 1/21/2015 Expires: 1/21/2016 [For RP ice Label Here] OPENED: 3/5/15 RCD Solution out of range. Leak-checked simulator and it was not sealing tight. Replaced simulator with str: VT 3. Performed temperature matching + solution change. RCD 3/5/15	R. Driscoll DMT TECH. Also trimmed tubing in sim. tower for tight fit on quick connects.
3/6/15	[For RPC or Simulator Solution Change Place Label Here] Pump Error x2 - Removed From Service	J. Lehmann DMT Supervisor
3/6/15	Tubing was kinked in simulator tower. Fixed tubing and ran an Accuracy + Precision check. RCD 3/6/15	R. Driscoll DMT TECH.
3/18/15	Possible rapid solution depletion. Checked new [For RPC or Simulator Solution Change Place Label Here] simulator for leaks, checked internal tubing + 5-way valve. No issues found. Ran ATP test with solution running at 0.096 g/210L (in range) RCD 3/18/15	R. Driscoll DMT TECH.

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
4/27/15	DataMaster Simulator Solution Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: 4/27/15 JBL [For RPC e Label Here]	J. Lehnen DMT Supervisor
	Solution change. Solution Out of Range .081 Average	
27 Apr 15	DataMaster Simulator Solution Lot # 15-05-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 2/19/2015 Expires: 2/19/2016 OPENED: 4/27/15 JBL [For RP ice Label Here]	J. Lehnen J. Roy DMT Supervisor
	Solution change .098 Average	
5/5/15	Solution out of range. DMT Supervisor tested [For RPC or Simulator Solution Change Place Label Here] after and the solution was within range. Checked simulator for leaks (none found) + ran AtL. Solution is still within +/- 5% tolerance. Rerouted instrument vent tube the instrument	R. Discoll DMT Tech
6-30-15	DataMaster Simulator Solution Lot # 15-05-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 2/19/2015 Expires: 2/19/2016 OPENED: 6-30-15 CMS [For RPC e Label Here]	
	RPC / Solution Change	

DATAMASTER DMT MAINTENANCE LOG

Colchester PD
Department/Agency

100157
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/15/15	Simulator out of range - [For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 15-05-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 2/19/2015 Expires: 2/19/2016 OPENED: 13 Aug 15 Solution changed 0915	D. Venkova DMT Supervisor
10/2/15	Completed APM protocol, Downloaded data [For RPC or Simulator Solution Change Place Label Here] to USB drive, uploaded new software, and installed new I/O PCB, Connected to DPS Server.	R. Driscoll DMT Tech.
11/1/15	RPC [For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 15-12-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/28/2015 Expires: 07/28/2016 OPENED: 11/1/15	J. Lehmann DMT Supervisor
11/25/15	DMT displaying "Busy, please wait" and [For RPC or Simulator Solution Change Place Label Here] does not appear to be talking to the DPS server. Rebooted the instrument, ran a check-in test, a 2-test DUI, and a diagnostic. All tests uploaded with no issue. Possible server connection problem within DPS.	R. Driscoll DMT Tech.

DATAMASTER DMT MAINTENANCE LOG

Colchester Police

Department/Agency

100157

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>11/26/15</i>	DataMaster Simulator Solution Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016 OPENED: <i>11/26/15</i> [For RPC or Simulator Solution Change Place Label Here] <i>Solution Change .098 Average Range</i>	<i>J. Lehmann DMT Supervisor</i>
<i>2-9-16</i>	DataMaster External Standard Solution Lot # 15-23-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 10/29/2015 Expires: 10/29/2016 OPENED: [For RPC or Simulator Solution Change Place Label Here] <i>Solution Change RPC</i>	<i>Clancy</i>
<i>2/10/16</i>	<i>Standard dev. during RPC was .0019 with a limit of 0.0020. Wanted to make sure the simulator was leak tight. Tested simulator for leaks and found one at the lid ring. Tightened jar and ran a ATP test St. Dev now 0.0004.</i> [For RPC or Simulator Solution Change Place Label Here]	<i>R. Disall DMT Tech.</i>
<i>2/11/16</i>	DataMaster External Standard Solution Lot # 15-25-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/30/2015 Expires: 11/30/2016 OPENED: [For RPC or Simulator Solution Change Place Label Here] <i>Solution Change</i>	<i>D. Lumbard</i>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2 July 16	DataMaster External Standard Solution Lot # 16-05-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 4/19/2016 Expires: 4/19/2017 [For RPC or Simulator Solution Change Place Label Here] Label Here] OPENED: <i>RPC</i>	D. Kerlan #168
10/17/16	DataMaster External Standard Solution Lot # 16-04-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 3/25/2016 Expires: 3/25/2017 [For RPC or Simulator Solution Change Place Label Here] ce Label Here] OPENED: <i>RPC & Check IN</i>	Lohmeier M193
10/21/16	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT Tech.
1-15-17	DataMaster External Standard Solution Lot # 16-04-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 3/25/2016 Expires: 3/25/2017 [For RPC or Simulator Solution Change Place Label Here] Label Here] OPENED: 01-15-17 C Jones <i>Solution Change</i>	C Jones M187

DATAMASTER DMT MAINTENANCE LOG

Colchester Police

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/24/17	DMT External Standard Solution Lot # 16-21-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 9/13/2016 Expires: 9/13/2017 [For RP OPENED: 2-24-17 CMS ice Label Here] RPC / Solution Change	CMS MIST
4/1/17	DMT Simulator Solution Lot # 17-16-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 04/03/2017 Expires: 04/03/2018 [For RP OPENED: 4/1/17 DJB ice Label Here] INSTALL AFTER SIMULATOR UPGRADE	D. PATLAK VFL
6-7-17	DMT Simulator Solution Lot # 17-16-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 04/03/2017 Expires: 04/03/2018 [For R OPENED: 6-7-17 CMS ice Label Here] RPC Solution Change	Clony 6-7-17
10-30-17	DMT Simulator Solution Lot # 17-18-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 05/02/2017 Expires: 05/02/2018 [For RP OPENED: 10-30-17 CMS ice Label Here] RPC - Solution Change	C Jones 10-30-17

DATAMASTER DMT MAINTENANCE LOG

Colchester PD
Department/Agency

100157
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/31/17	<u>Installed version 2.00 PIC chip.</u> [For RPC or Simulator Solution Change Place Label Here]	<u>R. Driscoll</u> VFL
2-28-18	DMT Simulator Solution Lot # 17-18-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 05/02/2017 Expires: 05/02/2018 [For RPC or Simulator Solution Change Place Label Here] OPENED: <u>2-28-18</u> <u>RPC Solution Change</u>	<u>C Jones</u> 181
4/9/18	<u>Completed APM Protocol.</u> [For RPC or Simulator Solution Change Place Label Here]	<u>R. Driscoll</u> VFL
6/28/18	<u>RPC Solution Change</u> DMT External Std Solution Lot: 18-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/28/18 Exp Date: 02/28/19 Opened: <u>6/28/18</u>	<u>J. Lehneman</u> M193
11-5-18	DMT External Std Solution Lot: 18-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/28/18 Exp Date: 02/28/19 [For RPC or Simulator Solution Change Place Label Here] Opened: <u>11-5-18</u> <u>RPC / Solution Change</u>	<u>C Jones</u> M181

Colchester Police

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

100157

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/28/19	DMT External Std Solution Lot: 18-05-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/21/18 Exp Date: 03/21/19 Opened: 2-28-19 VB [For RPC or Simulator Solution Change Place Label Here] RPC / Solution Change	V. Bitca M194 CPD
3/11/19	DMT External Std Solution Lot: 19-07-100 EtOH Certified Conc. 0.103 Range: 0.098 - 0.108 Prep Date: 5/31/19 Exp Date: 05/31/20 Opened: 6-26-19 [For RPC or Simulator Solution Change Place Label Here] RPC Solution Change	J. Dukette CVFL
6-26-19	DMT External Std Solution Lot: 19-07-100 EtOH Certified Conc. 0.103 Range: 0.098 - 0.108 Prep Date: 5/31/19 Exp Date: 05/31/20 Opened: 6-26-19 [For RPC or Simulator Solution Change Place Label Here] RPC Solution Change	M181 C Jones
10-19-19	DMT External Std Solution Lot: 19-07-100 EtOH Certified Conc. 0.103 Range: 0.098 - 0.108 Prep Date: 5/31/19 Exp Date: 05/31/20 Opened: 10-19-19 MSW [For RPC or Simulator Solution Change Place Label Here] RPC / Solution Change	M192 M West

DATAMASTER DMT MAINTENANCE LOG

Colchester PD

Department/Agency

100157

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/14/2020	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here] - 1 st attempt sim out of range - 2 nd attempt passed	S. Stratton (VFL)
2-21-2020	DMT External Std Solution Lot: 19-13-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 9/23/19 Exp Date: 09/23/20 Opened: 2-21-2020 cmj [For RPC or Simulator Solution Change Place Label Here] Solution change / RPC	C Jones 181
7-3-2020	DMT External Std Solution Lot: 20-03-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 1/29/20 Exp Date: 01/29/21 Opened: 7-3-2020 cmj [For RPC or Simulator Solution Change Place Label Here] Solution Change / RPC	C Jones 181
7/15/2020	Installed DMT software update [For RPC or Simulator Solution Change Place Label Here]	S Stratton (VFL)

DATAMASTER DMT MAINTENANCE LOG

Colchester PD.

Department/Agency

100157

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	DMT External Std Solution [For RP Lot: 20-14-100 EtOH Certified Conc. 0.100 ice Label Here] Range: 0.095 - 0.105 Prep Date: 9/22/20 Exp Date: 09/22/22 Opened: <i>10/28/2020 CMS</i> <i>Solution Change ←</i> <i>RPC ←</i>	<i>C. Jones</i> <i>181</i> <i>V. Bitca</i> <i>194</i>
<i>2/1/21</i>	Completed ARM [For RPC or Simulator Solution Change Place Label Here]	<i>J Dulake</i> <i>(LVFL)</i>
	DMT External Std Solution [For RP Lot: 20-14-100 EtOH Certified Conc. 0.100 ice Label Here] Range: 0.095 - 0.105 Prep Date: 9/22/20 Exp Date: 09/22/22 Opened: <i>02-04-2021 MSW</i> <i>Solution change.</i> <i>RPC</i>	<i>Det. M. West</i> <i>192.</i> <i>02-04-2021</i>
	DMT External Std Solution [For RP Lot: 20-15-100 EtOH Certified Conc. 0.100 lace Label Here] Range: 0.095 - 0.105 Prep Date: 11/20/20 Exp Date: 11/20/22 Opened: <i>6-11-21 CMS</i> <i>Solution Change</i> <i>RPC</i>	<i>6-10-21</i> <i>C Jones</i> <i>✓</i>

DATAMASTER DMT MAINTENANCE LOG

Colchester PD
Department/Agency

100157
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	DMT External Std Solution Lot: 20-15-100 EtOH Certified Conc. 0.100 [For RPC Range: 0.095 - 0.105 11-2-21 cm³ e Label Here] Prep Date: 11/20/20 Exp Date: 11/20/22 Opened: 11-02-21 CM RPC / Solution Change	11/2/2021 C Jones L
11/27/22	Completed ARM Protocol [For RPC or Simulator Solution Change Place Label Here]	J Dickett (VFL)
02-28-2022	DMT External Std Solution Lot: 21-12-100 EtOH Certified Conc. 0.099 Label Here] Range: 0.094 - 0.104 02-28-2022 Prep Date: 10/8/21 msh Exp Date: 10/08/23 Opened: RPC	Michael West 02-28-2022
7/3/22	[For RPC or Simulator Solution Change Place Label Here] DMT External Std Solution Lot: 22-01-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 1/4/22 Exp Date: 01/04/24 Opened: 7/3/22 VB RPC / Solution changed	Victor Bitca 7/3/22

DATAMASTER DMT MAINTENANCE LOG

Colchester PD

Department/Agency

100157

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/15/22 ↓	DMT External Std Solution [For RPC Lot: 21-12-100 EtOH Certified Conc. 0.099 ce Label Here] Range: 0.094 - 0.104 Prep Date: 10/8/21 Exp Date: 10/08/23 Opened: 11/15/22 VB RPC / Solution Change	V. Bitca ↓
1/24/23	Completed APM [For RPC or Simulator Solution Change Place Label Here]	J Dukette (VFC)
02-02-2023	DMT External Std Solution [For RF Lot: 22-14-100 EtOH Certified Conc. 0.098 ace Label Here] Prep Date: 10/5/22 Exp Date: 10/05/24 Opened: 02-02-2023 MSW RPC	M Wess
	DMT External Std Solution [For RPC Lot: 22-12-100 EtOH Certified Conc. 0.101 ce Label Here] Range: 0.096 - 0.106 Prep Date: 5/31/22 Exp Date: 05/31/24 Opened: 7/2/23 VB RPC + Solution Change	V. Bitca

DATAMASTER DMT MAINTENANCE LOG

Colchester PD
Department/Agency

160157
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	DMT External Std Solution Lot: 23-06-100 EtOH Certified Conc. 0.100 [For RPC <u>11/11/23 - 181 Jones</u> e Label Here] Prep Date: 5/1/23 Exp Date: 05/01/25 Opened:	
	<i>Solution Change</i> <i>RPC</i>	<i>Clones</i> <i>181</i>
<u>1/18/24</u>	<i>Completed APM protocol</i> [For RPC or Simulator Solution Change Place Label Here]	<i>S. Stratton</i> <i>(VFL)</i>
	DMT External Std Solution Lot: 23-10-100 EtOH Certified Conc. 0.099 [For RI <u>2/21/24</u> <i>181 Jones</i> ace Label Here] Prep Date: 11/28/23 Exp Date: 11/28/25 Opened:	
	<i>Solution Change</i> <i>RPC</i>	<i>Clones</i> <i>181</i>
	DMT External Std Solution Lot: 23-06-100 EtOH Certified Conc. 0.100 [For RPC <u>6/14/24</u> <i>181 Jones</i> e Label Here] Prep Date: 5/1/23 Exp Date: 05/01/25 Opened:	
	<i>Solution Change</i> <i>RPC</i>	<i>Clones</i> <i>181</i>

Colchester PD

100157

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	DMT Quality Control Solution [For RPC Lot: 24-07-100 EtOH Certified Conc. 0.100 e Label Here] Prep Date: 5/29/24 Exp Date: 05/29/26 Opened: 11-2-2024 VB Solution change / RPC	V. Bitca 194
1/15/25	Removed for Service [For RPC or Simulator Solution Change Place Label Here]	A. Bolduc
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	