VERMONT FORENSIC LABORATORY

PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
http://vfl.vermont.gov

DataMaster DMT Site Evaluation Checklist

Agency or Site Information

Name: Colchester Police Department

Address: 835 Blakely Road Colchester, VT 05446

Phone: (802) 264-5555

Fax: (802) 264-5550

Personnel Information

Head of Site: Charles Kirker

Current DataMaster Supervisor (s):

Sgt. James Roy

Pfc. Chris Jones

~~Proposed~~ Additional DataMaster Supervisor(s):

Pfc. Derrick Kendrew

Pfc. Jack Lehneman

(Proposed) DataMaster DMT Environment

Limited Controlled Access:	(Y)	N
Low Activity Area:	(Y)	N
Unlimited Access by Other Law Enforcement Personnel	(Y)	N
Access by VFL Technical Services	(Y)	N
Air Conditioned	(Y)	N
Room Temperature 65° - 85° F	(Y)	N
Adequate Ventilation	(Y)	N
Free of Cleaning Supplies	(Y)	N
Clean	(Y)	N

(Proposed) DataMaster DMT Connection(s)

Power: Min 15 Amp	(Y)	N
Limited low power sharing:	(Y)	N

Phone Line Connection Number: ___-___-___

Phone Line Connection Type ___

(Proposed) DataMaster DMT Site Maintenance Agreement

Will notify VFL Alcohol Program staff of maintenance and supply needs promptly:	(Y)	N
Will notify VFL Alcohol Program staff of changes in building conditions or moves promptly:	(Y)	N
Will notify VFL Alcohol Program staff of relevant personnel changes at site promptly:	(Y)	N

VFL Representative:	Robert Driscoll (Print Name)	(Signature)	09/12/2012 (Date)
Agency Representative:	James Roy (Print Name)	(Signature)	12Sep12 (Date)