

DataMaster Monthly Update Check-list

Department/Agency: Colchester PD DM serial #: 100157

Date: 5 15 115 Log for April

	Provided	None this month
Operators Logbook	☒	☐
Maintenance Logbook	☒	☐
Breath Summary Report	☒	☐
Status Summary Report	☒	☐
Calibration	☐	☒
Certification	☐	☒
Installation	☐	☒
Annual Preventative Maintenance	☐	☒
Routine Performance Check	☐	☒
Simulator Solution Change	☐	☒
Other Reports	☐	☐

List other report type and date:

DATAMASTER DMT OPERATOR USE LOG

Colchester PD
Department/Agency

100157
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
3/5/15	R. Driscoll	TEST / solution change	.098	—
03/06/15	Lehnehan	Jagves	Pump Error	Pump Error
3/6/15	out of Service	Lehnehan J	—	—
3/6/15	R. Driscoll	A+P check	.097	—
3/7/15	J. Sawyer	Atfred Jagves	.098	—
3/14/15	C. Mellon	Leonard Parks	.096	—
3/18/15	J. WYSKIEL	LEONARDO DASILVA	.096	.094
3/18/15	J. WYSKIEL	LEONARDO DASILVA	SCREEN WENT TO NPAS	
3/18/15	Lehnehan	Test	.096	
3/18/15	R. Driscoll	TEST / A+P		—
3/21/15	Jones Gregory Moehn	Gregory Moehn	.046	
3/27/15	J. TREIER	DANIEL INFANTES	.095	
3/31/15	J. TREIER	PATRICK ALWARD	.095	.094
4/1/15	Brunette	Ryan J. Ward	.095	.095
4/2/15	Lehnehan	Leigh-Ann Francis	.094	
4-26-15	FISH	Nicholas G. Ford	.094	
4/26/15	Lehnehan	Robert E Gorruso	.094	
4/26/15	TREIER	MATTHEW G. LOGAN	.094	—
4/27/15	J. WYSKIEL	CHRISTOPHER MESSER	.094 .098	.093
4/27/15	Lehnehan	Solution Change	.081	

Colchester Police
Department/Agency

[illegible]

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
4/27/15	DataMaster Simulator Solution Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: 4/27/15 JBL [For RPC or Simulator Solution Change Place Label Here]	J. Lehnen DMT Supervisor
	Solution change. Solution Out of Range .081 Average	
27 Apr 15	DataMaster Simulator Solution Lot # 15-05-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 2/19/2015 Expires: 2/19/2016 OPENED: 4/27/15 JBL [For RPC or Simulator Solution Change Place Label Here]	J. Lehnen J. Roy DMT Supervisor
	Solution change .098 Average	
5/5/15	Solution out of range. DMT Supervisor tested [For RPC or Simulator Solution Change Place Label Here] after and the solution was within range. Checked simulator for leaks (none found) + ran A+P. Solution is still within +/- 5% tolerance. Rerouted instrument vent tube from the bottom to the back of the instrument	R. Disell DMT Tech
	[For RPC or Simulator Solution Change Place Label Here]	

Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
04/11/2015	03:51:39	SBPD	D BRUNELLE	0.103	0.099	0.096	0.095	
04/21/2015	00:20:44	COLCHESTER PD	JACK LEHNEMAN	0.074		0.094		
04/26/2015	01:54:28	COLCHESTER PD	MICHAEL FISH	0.151	0.163	0.094	0.094	
04/26/2015	04:44:57	COLCHESTER PD	JACK LEHNEMAN	0.187		0.094		
04/26/2015	21:58:10	COLCHESTER PD	JESSE L TREIER	0.116		0.094		
04/27/2015	04:27:12	COLCHESTER PD	JEREMY A WYSKIEL	0.206	0.207	0.094	0.093	

Page 1
05/05/2015

DMT Serial Number #100157 Page 1 of 1 05/05/2015 8:31 AM