

DataMaster Monthly Update Check-list

Department/Agency: Colchester PD DM serial #: 100157

Date: 6/10/2015

	Provided	None this month -
Operators Logbook	☒	☐
Maintenance Logbook	☒	☐
Breath Summary Report	☒	☐
Status Summary Report	☒	☐
Calibration	☐	☒
Certification	☐	☒
Installation	☐	☒
Annual Preventative Maintenance	☐	☒
Routine Performance Check	☐	☒
Simulator Solution Change	☐	☒
Other Reports	☐	☒

List other report type and date:

DATA MASTER DMT OPERATOR USE LOG

Colchester Police
Department/Agency

100157
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
4/27/15	Roy	Sootam change	.098	
5/2/15	Dewey	Peaslee, D	.1093	
5/2/15	Lehnenman	A+P Test	.1096	Orange
5/2/15	Lehnenman	Test check in	.097	
5/5/15	R. Driscoll	TEST /A+P	.096	—
5/5/15	R. Driscoll	TEST /A+P	.097	—
5/10/15	FIS L	Jennifer Bussanotte	.096	
5/10/15	WYSKIEL	BUSHEY, MERISSA	.096	.095
5/17/15	Fish	Brown, Shawn	.096	.095
5/22/15	GUTIERREZ	HULBART, SHAB	.096	
5/23/15	Lehnenman	Pilon, Michael	.1096	

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
4/27/15	DataMaster Simulator Solution Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: 4/27/15 JBL [For RPC e Label Here]	J. Lehnenman DMT Supervisor
	Solution change. Solution Out of Range .081 Average	
27 Apr 15	DataMaster Simulator Solution Lot # 15-05-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 2/19/2015 Expires: 2/19/2016 OPENED: 4/27/15 JBL [For RP ice Label Here]	J. Lehnenman J. Roy DMT Supervisors
	Solution change .098 Average	
5/5/15	Solution out of range. DMT Supervisor tested [For RPC or Simulator Solution Change Place Label Here] after and the solution was within range. Checked simulator for leaks (none found) + ran A+B. Solution is still within +/- 5% tolerance. Rerouted instrument vent tube from the bottom to the back of the instrument	R. Disc. II DMT Tech
	[For RPC or Simulator Solution Change Place Label Here]	

Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
05/02/2015	00:08:42	COLCHESTER PD	D C DEMEY	REFUSED		0.094		
05/02/2015	00:28:42	COLCHESTER PD	D C DEMEY			0.092		Simulator Out of Range
05/02/2015	00:32:17	COLCHESTER PD	D C DEMEY			0.093		Simulator Out of Range
05/02/2015	00:35:56	COLCHESTER PD	D C DEMEY			0.093		Simulator Out of Range
05/02/2015	02:14:41	COLCHESTER PD	JACK LEHNEMAN	0.000		0.097		
05/10/2015	03:01:40	COLCHESTER PD	MICHAEL FISH	REFUSED		0.096		
05/10/2015	18:40:34	COLCHESTER PD	JEREMY A WISKIETL	0.245	0.241	0.096	0.095	
05/16/2015	01:28:39	COLCHESTER PD	D C DEMEY	REFUSED		0.096		
05/17/2015	14:42:30	COLCHESTER PD	MICHAEL FISH	0.118	0.119	0.096	0.095	
05/22/2015	05:46:24	COLCHESTER PD	STEPHEN GUTIERREZ	REFUSED		0.096		
05/23/2015	04:12:36	COLCHESTER PD	JACK LEHNEMAN	REFUSED		0.096		

Status Record History