

SOLUTION CHANGE

DataMaster DMT: 100157
Location: Colchester PD
Date: 11/26/2015
Performed by: JACK B LEHNEMAN



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-13-100
Range = 0.094 - 0.104
Average = 0.098 g/210L
Std Dev = 0.0005

Revised 11/20/15

Simulator Temperature: 34.0°C

Performed by

J. B. Lehman

Date

11/26/2015