

Department/Agency: Colchester PD DM serial #: 100157

Date: 6 10 1 2016 - 6-30-16

	Provided	None this month -
Operators Logbook	☒	☐
Maintenance Logbook	☐	☒
Breath Summary Report	☒	☐
Status Summary Report	☒	☐
Calibration	☐	☒
Certification	☐	☒
Installation	☐	☒
Annual Preventative Maintenance	☐	☒
Routine Performance Check	☐	☒
Simulator Solution Change	☐	☒
Other Reports	☐	☒

List other report type and date:

DATAMASTER DMT OPERATOR USE LOG

Colchester Police

Department/Agency

100157

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
4-12-16	AHERLUND	SAMUEL L. HECA	.096	—
4-13-16	Schaffer	William Murray	.097	
4/17/16	GUTIERREZ	WILLIAM MURRAY	.095	
4/22/16	MENEN	David Reuschel	.095	
4/29/16	Wyskiel	Henry Emery	.097	
4/30/16	GUTIERREZ	ALLYSON CLODRECHER	.097	
5-1-16	Hendry MPD	Destiny Little	.094	
05/04/16	West	Justine Siple	.095	
5/7/16	Bressler	John M. Miller	.094	
05/11/16	West	Nancy Zeno	.095	
5/14/16	Bitca	Kristin Dahlen	.097	
5/15/16	Wyskia	Troy Berland	.094	
5-21-16	Fish	Robert ST. Peter		
5-24-16	Wyskia	Robert J. Butts	.093	out of Range
5-24-16	Wyskia (Acon & Daropos)	Robert J. Butts	.093	
24 May 16	Test / Solution change (ack)	—	.098	
3 June 16	Lehman Descoteaux Jose	Descoteaux, Joseph	.098	
6-5-16	SM. UEL Fish Shannon	Shannon Plice	.097	
6-13-16	V. Bitca	Dylan Harris	.097	
06/25/2016	West	David Kost	.1097	

Status Record History

Status Date	Status Time	Test Type	Status Message
06/25/2016	14:53:54	DUI Subject Record	Simulator Out of Range