

Department/Agency: Colchester Police DM serial #: 100157

Date: 5101116 - 5/31/16 4/8/16

	Provided	None this month -
Operators Logbook	☒	☐
Maintenance Logbook	☒	☐
Breath Summary Report	☒	☐
Status Summary Report	☒	☐
Calibration	☐	☒
Certification	☐	☒
Installation	☐	☒
Annual Preventative Maintenance	☐	☒
Routine Performance Check	☐	☒
Simulator Solution Change	☒	☐
Other Reports	☐	☐

List other report type and date:

DATA MASTER DMT OPERATOR USE LOG

COLCHESTER POLICE
Department/Agency

100157
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
4-12-16	AKERLIN	SAMUEL L. HECA	.096	—
4-13-16	Schuffor	William Murray	.097	
4/17/16	GUTIERREZ	WILLIAM M. RAY	.095	
4/22/16	MENEN	David Reuschel	.095	
4/29/16	Wyskiel	Henry Emery	.097	
4/30/16	GUTIERREZ	ALLYSON CLODRECHT	.097	
5-1-16	Hendry MPD	Destiny Little	.094	
05/04/16	West	Justine Siple	.095	
5/7/16	Bressler	John M. Miller	.094	
05/11/16	West	Nancy Zeno	.095	
5/14/16	Bitca	Kristin Dahlen	.097	
5/15/16	Wyskiel	Troy Beland	.094	
5-21-16	Fish	Robert ST. Peter		
5-24-16	Wyskiel	Robert J. Butts	.093	out of Range
5-24-16	Wyskiel (Acon & Daropow)	Robert J. Butts	.093	
24 May 16	Test / Solomon change Lehman	—	.098	
3 June 16	Descoteaux Joe	Descoteaux, Joseph	.098	
6-5-16	M. U. Fish Shannon Place	Shannon Place	.097	

DATA MASTER DMT MAINTENANCE LOG

Colchester Police
Department/Agency

100157
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>11/26/15</i>	<u>DataMaster Simulator Solution</u> Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016 OPENED: <i>11/26/15</i> [For RPC] <i>Solution Change .098 Average Range</i>	<i>J. Lehmann DMT Supervisor</i>
<i>2-9-16</i>	<u>DataMaster External Standard Solution</u> Lot # 15-23-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 10/29/2015 Expires: 10/29/2016 OPENED: [For] <i>Solution Change RPC</i>	<i>Clay</i>
<i>2/10/16</i>	<i>Standard dev. during RPC was .0019 with a limit</i> [For RPC or Simulator Solution Change Place Label Here] <i>of 0.0020. Wanted to make sure the simulator was</i> <i>leak tight. Tested simulator for leaks and found one</i> <i>at the lid ring. Tightened jar and ran a ATP test</i> <i>St. Dev now 0.0004.</i>	<i>R. Disall DMT Tech.</i>
<i>2/11/16</i>	<u>DataMaster External Standard Solution</u> Lot # 15-25-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/30/2015 Expires: 11/30/2016 OPENED: [For RF] <i>Solution Change</i>	<i>D. Vanden</i>

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
05/01/2016	04:11:20	MILTON PD	NICK HENDRY	0.078		0.094		
05/04/2016	00:43:55	COLCHESTER PD	MICHAEL S WEST	0.190		0.095		
05/07/2016	18:21:25	COLCHESTER PD	JAIME L BRESSLER	0.087	0.097	0.094	0.093	
05/11/2016	21:15:11	COLCHESTER PD	MICHAEL S WEST	0.205		0.095		
05/14/2016	19:37:59	COLCHESTER PD	VICTOR BITCA	0.145		0.095		
05/15/2016	05:20:27	COLCHESTER PD	JEREMY A WYSKIEL	REFUSED		0.094		
05/21/2016	17:38:07	COLCHESTER PD	MICHAEL FISH	0.155	0.158	0.094	0.093	
05/24/2016	02:04:11	COLCHESTER PD	JEREMY A WYSKIEL	0.143		0.093	0.090	Simulator Out of Range
05/24/2016	02:14:12	COLCHESTER PD	JEREMY A WYSKIEL	0.144		0.093		
06/03/2016	20:40:33	COLCHESTER PD	JACK B LEHNEVAN	REFUSED		0.098		
06/05/2016	04:18:40	COLCHESTER PD	MICHAEL FISH	REFUSED		0.097		
06/05/2016	04:39:57	COLCHESTER PD	MICHAEL FISH	0.212		0.097		

Status Record History

[illegible]

SOLUTION CHANGE

DataMaster DMT: 100157
Location: Colchester PD
Date: 05/24/2016
Performed by: DERRICK C KENDREW



Accuracy and Precision Check

Concentration	= 0.100 g/210L
Lot #	= 15-25-100
Range	= 0.095 - 0.105
Average	= 0.098 g/210L
Std Dev	= 0.0005

Simulator Temperature: 34.0°C

Performed by

A handwritten signature in black ink, appearing to read "DK", is written over a horizontal line.

Date

05/24/2016