

SOLUTION CHANGE

DataMaster DMT: 100157
Location: Colchester PD
Date: 05/24/2016
Performed by: DERRICK C KENDREW



Accuracy and Precision Check

Concentration = 0.100 g/210L
Lot # = 15-25-100
Range = 0.095 - 0.105
Average = 0.098 g/210L
Std Dev = 0.0005

RCD
5/25/16

Simulator Temperature: 34.0°C

Performed by

A handwritten signature in black ink, appearing to be "DCK", is written over a horizontal line.

Date

05/24/2016