

Datamaster Monthly Update Checklist

Department/Agency: Colchester PD DM serial #: 100157

Date: 12/01/17 - 12/31/17 Sent: 1/5/18

	Provided	None this month
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Breath Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Status Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Calibration	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Certification	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Installation	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Annual Preventative Maintenance	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Routine Performance Check	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Simulator Solution Change	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Other Reports	<input checked="" type="checkbox"/>	<input type="checkbox"/>

List other report type and date:
Diagnostic Results

CPS

DATAMASTER DMT OPERATOR USE LOG

Department/Agency

100157

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
7-2-17	V. BITCA	Thelma Trayah	0.100%	
7/4/17	J. Baur	James White	.098	
7/5/17	J. Baur	Michael Burnham	.098	
7/5/17	J. Baur	Michael Burnham	.098	0.098 SAB
7/5/17	Z. Smith	Theresa Granger	.099	
7/9/17	S. GUTIERREZ	SHELBY REPOSA	.101	
07-21-17	M. West	Alfred Hotescheiter	.099	.098
8/2/17	Z. Smith	Karin Couillard	.099	.098
8/17/17	Z. Smith	James White	0.098 .097	
8/20/17	S. GUTIERREZ	JUSTICE SWEENEY	OUT OF RANGE	
09/02/17	F. Gonyaw	Walter Pahlca	.099	.098
9/27/17	F. Gonyaw	Danielle Leclair	.098	.097
9/29/17	S. GUTIERREZ	DAVID BABIS	.098 .101	
10/1/17	V. BITCA	Mark Adams	0.98	
10/10/17	S. GUTIERREZ	KERVIN DESMOND	.098	.097
10-31-17	C Jones	RPC	.101	
11-05-17	C Jones	Munzo, Ryan	.100	
11-19-17	V. BITCA	Shannon Masou	0.99	
12/2/17	Z. Smith	Lynn Norman	0.99	
12/10/17	S. GUTIERREZ	LYON, JAMES	.099	.099

Status Record History

Status Date	Status Time	Test Type	Status Message

DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location: COLCHESTER PD
Calibration Date: 04/11/2017
Certification Date: 04/11/2017
Installation Date: 04/11/2017
Test Date: 12/15/2017
Test Time: 10:00:02



VERSIONS

DMT: 2.03
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.9°C
Breath Tube = 48.0°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.66 V
Cooler Voltage = 1.67 V
Bias Voltage = 80 V
Chopper Freq = 541 Hz

PUMP INFO

Flow Rate = 3.841 L/M

DETECTOR INFO

	ON	OFF
MAX (V)	0.0169	0.0207
MIN (V)	0.0154	0.0187

FILTER INFO

Filter 1	0.019	Zero = true
Filter 2	0.317	Zero = true
Filter 3	1.431	Zero = true

CALIBRATION CHECK

Xq = 0.095 1.03%