

Calibrator Monthly Update Checklist

Department/Agency: Colchester P.D. DM serial #: 100157

Date: 1/10/2017 - 1/31/2017 2/15/17

	Provided	None this month
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Breath Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Status Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Calibration	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Certification	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Installation	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Annual Preventative Maintenance	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Routine Performance Check	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Simulator Solution Change	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Other Reports	<input type="checkbox"/>	<input checked="" type="checkbox"/>

List other report type and date:

DATAMASTER DMT OPERATOR USE LOG

Colchester PD

Department/Agency

100157

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
10/30/16	J. TREZER	TANYA SANDER		
11-1-16	C Jones	Shane Gallo	.096	.096
11/3/16	S. GUTIERREZ	DEBORAH MCCUSKER	.096	
11/13/16	J. TREZER	TASHA MANCHESTER	.096	.094
11/13/16	S. GUTIERREZ	NELSON CARAGNON	.096	.095
11/30/16	V. BITCA	Melissa Bouchard	.095	
12/02/16	M. AKERLEND	JANAK GURUNG	.095	
12-08-16	West	Charles Cota	.095	
12-20-16	V. BITCA	Samuel Bombard	.094	
01-06-17	West	Matthew Verbeent	.094	.093
01-19-17	Jones	Sgt. Chunn	.098	
01/20/17	Fontana	Sgt. Mercure	.096	.094
1/29/17	R. CURRIER	JESSE ILGES		
2/1/17	V. BITCA	Robert Lopez	.093	
2-5-17	C Jones	Charles Nahn	.098	
2-6-17	C Jones	Gary Pasquale	.095	

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/24/16	<p>DataMaster External Standard Solution Lot # 16-05-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 4/19/2016 Expires: 4/19/2017</p> <p>[For RPC c Label Here]</p> <p>OPENED:</p> <p style="font-size: 2em; text-align: center;"><i>RPC</i></p>	D. Kerlow #168
10/17/16	<p>DataMaster External Standard Solution Lot # 16-04-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 3/25/2016 Expires: 3/25/2017</p> <p>[For RP ce Label Here]</p> <p>OPENED:</p> <p style="font-size: 2em; text-align: center;"><i>RPC & check ID</i></p>	Lohmann M193
10/21/16	<p style="text-align: center;">Completed APM Protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	R. Driscoll DMT Tech.
1-15-17	<p>DataMaster External Standard Solution Lot # 16-04-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 3/25/2016 Expires: 3/25/2017</p> <p>[For RPC Label Here]</p> <p>OPENED: 01-15-17 C Jones</p> <p style="font-size: 2em; text-align: center;"><i>Solution Change</i></p>	C Jones M187

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
01/06/2017	23:21:20	COLCHESTER PD	MICHAEL S WEST	0.117		0.094	0.092	Simulator Out of Range
01/06/2017	23:30:04	COLCHESTER PD	MICHAEL S WEST	0.114		0.093	0.092	RESPONSE TIMEOUT
01/14/2017	00:14:16	COLCHESTER PD	FRANCIS C GONYAW	0.158		0.093	0.092	Simulator Out of Range
01/14/2017	00:21:33	COLCHESTER PD	FRANCIS C GONYAW	0.161		0.093	0.092	Simulator Out of Range
01/14/2017	00:28:54	COLCHESTER PD	FRANCIS C GONYAW	0.150		0.093	0.092	Simulator Out of Range
01/22/2017	10:25:50	COLCHESTER PD	JEFFREY L FONTAINE	0.195	0.191	0.096	0.094	Simulator Out of Range
01/29/2017	14:05:00	FISH & WILDLIFE	ROBERT L CURRIER			0.095		RESPONSE TIMEOUT
01/29/2017	14:37:34	FISH & WILDLIFE	ROBERT L CURRIER	0.063	0.064	0.094	0.094	

SOLUTION CHANGE

DataMaster DMT: 100157
Location: Colchester PD
Date: 01/15/2017
Performed by: CHRISTOPHER M JONES



Accuracy and Precision Check

Concentration = 0.098 g/210L
Lot # = 16-04-100
Range = 0.093 - 0.103
Average = 0.098 g/210L
Std Dev = 0.0007

Simulator Temperature: 34.0°C

Performed by

 C Jones 181

Date

 01/15/2017