

Datamaster Monthly Update Checklist

Department/Agency: Colchester Police DM serial #: 100157

Date: 8/15/2017 for 7/1/17 - 7/31/17

	Provided	None this month
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Breath Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Status Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Calibration	<input type="checkbox"/>	<input type="checkbox"/>
Certification	<input type="checkbox"/>	<input type="checkbox"/>
Installation	<input type="checkbox"/>	<input type="checkbox"/>
Annual Preventative Maintenance	<input type="checkbox"/>	<input type="checkbox"/>
Routine Performance Check	<input type="checkbox"/>	<input type="checkbox"/>
Simulator Solution Change	<input type="checkbox"/>	<input type="checkbox"/>
Other Reports	<input checked="" type="checkbox"/>	<input type="checkbox"/>

List other report type and date:
Diagnostic Result

DATAMASTER DMT OPERATOR USE LOG

C.P.D.

Department/Agency

100157

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
3/8/17	Locke, Paul	Heidi Cass	.097	
3/9/17	Coulombe, Jaseo	Brandon E Scarfone	.098	.096
4/2/17	GUTIERREZ, STEPHEN	KIEVIN DIDEMBAKI	.098	
4/4/17	Jones, Chris	Evan Harwood	.098	
4/7/17	Roy, James	Jane Doe	.096	
4/7/17	Lehmann, Jack John Doe	John Doe	.098	
4/11/17	DAVID ATAK	TEST/INSTALL	0.099	
4/13/17	V. BITCA	Christopher Mason	0.099	
4-14-17	C Jones	Chad Hotchkiss	.090	
4/17/17	J. Bressler	Arthur Burnier	.099	
04-21-17	M West	Nicholas Deannaops	.099	
4/22/17	V BITCA	Anthony Brigante	.099	.098
5/3/17	FISH	Alexa Craft	.101	
05-25-17	West	Kelley Devins	.1099	.097
6-7-17	Jones	RPC	.100	
6-9-17	FISH	Ashley Traph	.1099	
6-10-17	Polandini	HARVEY, LILLIAN	.101	
6-15-17	V. BITCA	Moran, Kelly	.100	
6/16/17	R. CURRIER	TERIN K PORTANTO	.08	
7-1-17	C Jones	Amanda Reilly	.098	

DECATUR RESULTS SUMMARY

08/16/2017

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
07/02/2017	03:23:03	COLCHESTER PD	CHRISTOPHER M JONES	0.079		0.098		RESPONSE TIMEOUT
07/02/2017	23:24:46	COLCHESTER PD	VICTOR BITCA	0.131		0.100		
07/04/2017	21:27:55	COLCHESTER PD	JEFFREY A BAUR	REFUSED		0.098		
07/05/2017	01:08:44	COLCHESTER PD	JEFFREY A BAUR	REFUSED		0.098		
07/05/2017	01:30:38	COLCHESTER PD	JEFFREY A BAUR	0.084		0.098		
07/05/2017	18:53:11	COLCHESTER PD	ZACHARY M SMITH	0.133		0.099		
07/09/2017	02:10:02	COLCHESTER PD	STEPHEN J GUTIERREZ	0.213		0.099		
07/21/2017	09:44:41	COLCHESTER PD	MICHAEL S WEST	0.142	0.140	0.099	0.098	

DIAGNOSTIC RESULT



DataMaster DMT: 100157
Location: COLCHESTER PD
Calibration Date: 04/11/2017
Certification Date: 04/11/2017
Installation Date: 04/11/2017
Test Date: 08/15/2017
Test Time: 10:00:03

VERSIONS

DMT: 2.03
PIC: 2.05
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.9°C
Breath Tube = 45.3°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.66 V
Cooler Voltage = 1.67 V
Bias Voltage = 80 V
Chopper Freq = 542 Hz

PUMP INFO

Flow Rate = 3.966 L/M

DETECTOR INFO

	ON	OFF
MAX (V)	-0.0079	-0.0051
MIN (V)	-0.0098	-0.0069

FILTER INFO

Filter 1	-0.007	Zero = true
Filter 2	0.296	Zero = true
Filter 3	1.405	Zero = true

CALIBRATION CHECK

Xq = 0.094 0.19%