

Datamaster Monthly Update Checklist

Department/Agency: Colchester Police DM serial #: 100157

Date: 4 117 117 3-01-17 thru 3-31-17

	Provided	None this month
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Breath Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Status Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Calibration	<input type="checkbox"/>	<input type="checkbox"/>
Certification	<input type="checkbox"/>	<input type="checkbox"/>
Installation	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Annual Preventative Maintenance	<input type="checkbox"/>	<input type="checkbox"/>
Routine Performance Check	<input type="checkbox"/>	<input type="checkbox"/>
Simulator Solution Change	<input type="checkbox"/>	<input type="checkbox"/>
Other Reports	<input checked="" type="checkbox"/>	<input type="checkbox"/>

List other report type and date:

Diagnostic Results

DATAMASTER DMT MAINTENANCE LOG

Colchester Police
Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/24/17	<p>DMT External Standard Solution Lot # 16-21-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 9/13/2016 Expires: 9/13/2017 [For RP] OPENED: 2-24-17 CMS [Place Label Here]</p> <p style="font-size: 1.2em; text-align: center;">RPC / SOLUTION CHANGE</p>	<p style="font-size: 1.5em;">CMS</p> <p style="font-size: 1.5em;">MIS</p>
4/11/17	<p>DMT Simulator Solution Lot # 17-16-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 04/03/2017 Expires: 04/03/2018 [For RP] OPENED: 4/11/17 DJP [Place Label Here]</p> <p style="font-size: 1.2em; text-align: center;">INSTALL AFTER SIMULATOR UPGRADE</p>	<p style="font-size: 1.5em;">D. PATLAK</p> <p style="font-size: 1.5em;">VFL</p>
[Faint handwritten text]	<p style="text-align: center;">[For RPC or Simulator Solution Change Place Label Here]</p>	
	<p style="text-align: center;">[For RPC or Simulator Solution Change Place Label Here]</p>	

PAULI NEVADA HILLS

04/17/2017

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
03/08/2017	15:04:48	MILTON PD	PAUL LOCKE	0.224		0.097		
03/09/2017	02:19:56	MILTON PD	WARREN G. COULOMBE	0.156	0.166	0.098	0.096	

DMT Serial Number #100157

Status Record History

Status Date	Status Time	Test Type	Status Message

INSTALLATION REPORT

DataMaster DMT: 100157
Location: COLCHESTER PD
Calibration Date: 04/11/2017
Certification Date: 04/11/2017
Installation Date: 04/11/2017
Installed by: DAVID PATLAK



Default Options Loaded

Diagnostic Results

VERSIONS
DMT: 2.03
PIC: 2.05
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.9°C
Breath Tube = 45.4°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.66 V
Cooler Voltage = 1.67 V
Bias Voltage = 80 V
Chopper Freq = 531 Hz

PUMP INFO

Flow Rate = 3.850 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX (V)	-0.0303	-0.0266
MIN (V)	-0.0323	-0.0289

FILTER INFO

Filter 1	-0.028	Zero = true
Filter 2	0.272	Zero = true
Filter 3	1.369	Zero = true

CALIBRATION CHECK

Xq = 0.093 0.37%

INSTALLATION PASSED

Accuracy and Precision Check


Concentration	= 0.101 g/210L
Lot #	= 17-16-100
Range	= 0.096 - 0.106
Average	= 0.099 g/210L
Std Dev	= 0.0004

RF Detection Test

Passed

Sample Acceptance Test

Passed



Performed by



Date

04/11/2017

Reviewed by

Date

DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location: COLCHESTER PD
Calibration Date: 04/11/2017
Certification Date: 04/11/2017
Installation Date: 04/11/2017
Test Date: 04/15/2017
Test Time: 10:00:02



VERSIONS

DMT: 2.03
PIC: 2.05
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 49.1°C
Breath Tube = 44.3°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.66 V
Cooler Voltage = 1.67 V
Bias Voltage = 80 V
Chopper Freq = 539 Hz

PUMP INFO

Flow Rate = 3.985 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX(V)	-0.1484	-0.1445
MIN(V)	-0.1500	-0.1461

FILTER INFO

Filter 1	-0.146	Zero = true
Filter 2	0.164	Zero = true
Filter 3	1.300	Zero = true

CALIBRATION CHECK

Xq = 0.094 0.47%