

Datamaster Monthly Update Check-list

Department/Agency: Colchester PD DM serial #: 100157

Date: 11/05/17 to 11/30/17 12/8/17

	Provided	None this month
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Breath Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Status Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Calibration	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Certification	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Installation	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Annual Preventative Maintenance	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Routine Performance Check	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Simulator Solution Change	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Other Reports	<input checked="" type="checkbox"/>	<input type="checkbox"/>

List other report type and date:

Diagnostic Result

CPD

DATAMASTER DMT OPERATOR USE LOG

100157

Department/Agency

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
7-2-17	V. BITCA	Thelma Trayah	0.100%	
7/4/17	J. Baur	James White	.098	
7/5/17	J. Baur	Michael Burnham	.098	
7/5/17	J. Baur	Michael Burnham	.098	0.094 SAB
7/5/17	Z. Smith	Theresa Granger	.1099	
7/9/17	S. GUTIERREZ	SHELBY REPOSA	.101	
07-21-17	M. West	Alfred Holzschetter	.099	.1098
8/2/17	Z. Smith	Karin Couillard	.099	.098
8/17/17	Z. Smith	James White	0.099 .097	
8/20/17	S. GUTIERREZ	JUSTIN SWEENEY	OUT OF RANGE	
09/02/17	F. Gonyaw	Walter Pahlco	.099	.098
9/27/17	F. Gonyaw	Danielle Leclair	.098	.097
9/29/17	S. GUTIERREZ	DAVID BABIS	.098 .101	
10/1/17	V. BITCA	Mark Adams	0.98	
10/10/17	S. GUTIERREZ	KEVIN DESMAREZ	.098	.097
10-21/17	C Jones	RPC	.101	
11-05-17	C Jones	Munzo, Ryan	.100	
11-19-17	V. BITCA	Shannon Mason	0.99	
12/2/17	Z. Smith	Lynn Norman	0.99	

DATAMASTER DMT MAINTENANCE LOG

Calchester PD
Department/Agency

100157
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/31/17	Installed version 2.00 PIC chip. [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFL
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

DLTGAUU NEUUTU ILLSULTY

12/08/2017

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
11/19/2017	13:50:48	COLCHESTER, PD	VICTOR, BITCA	0.228		0.099		

Status Record History

Status Date	Status Time	Test Type	Status Message

DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location: COLCHESTER PD
Calibration Date: 04/11/2017
Certification Date: 04/11/2017
Installation Date: 04/11/2017
Test Date: 11/15/2017
Test Time: 10:00:02



VERSIONS

DMT: 2.03
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.8°C
Breath Tube = 45.3°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.66 V
Cooler Voltage = 1.67 V
Bias Voltage = 80 V
Chopper Freq = 541 Hz

PUMP INFO

Flow Rate = 4.056 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX (V)	-0.0045	-0.0006
MIN (V)	-0.0061	-0.0024

FILTER INFO

Filter 1	-0.002	Zero = true
Filter 2	0.298	Zero = true
Filter 3	1.403	Zero = true

CALIBRATION CHECK

Xq = 0.094 0.19%