

SOLUTION CHANGE

DataMaster DMT: 100157
Location: Colchester PD
Date: 01/15/2017
Performed by: CHRISTOPHER M JONES



Accuracy and Precision Check

Concentration = 0.098 g/210L
Lot # = 16-04-100
Range = 0.093 - 0.103
Average = 0.098 g/210L
Std Dev = 0.0007

ROS 1/17/17

Simulator Temperature: 34.0°C

Performed by C Jones 181

Date 01/15/2017