

ANNUAL PREVENTATIVE MAINTENANCE



DataMaster DMT: 100157
Location: COLCHESTER PD
Calibration Date: 04/11/2017
Certification Date: 04/11/2017
Installation Date: 04/11/2017
APM Date: 04/09/2018
Performed by: ROBERT DRISCOLL

Diagnostic Results

VERSIONS
DMT: 2.03
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 49.4°C
Breath Tube = 42.8°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.66 V
Cooler Voltage = 1.67 V
Bias Voltage = 80 V
Chopper Freq = 533 Hz

PUMP INFO

Flow Rate = 4.059 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX (V)	-0.0024	0.0018
MIN (V)	-0.0041	0.0000

FILTER INFO

Filter 1	0.001	Zero = true
Filter 2	0.300	Zero = true
Filter 3	1.396	Zero = true

CALIBRATION CHECK

Xq = 0.093 0.37%

APM PASSED

Simulator Vapor Test


Concentration = 0.100 g/210L
Lot # = 17-18-100
Range = 0.095 - 0.105
Simulator Vapor = 0.099 g/210L

RF Detection Test

Passed

Sample Acceptance Test

Passed



Performed by

RD

Date

04/09/2018

Reviewed by

RD

Date

4/13/18

Simulator Temperature Check

Simulator Serial Number: MP3508

DMT Serial Number / Simulator Location: 100157 / Colchester PD

NIST Traceable Thermometer Serial Number: 160294076

As Found

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator	33.95	34.00	0.05	Y


* 34.0 +/- 0.2 degree Celsius acceptable range

Was temperature adjustment performed? Y / (N)

As Left (if applicable)

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator				

Notes:

Check Performed By: 

Date: 4/9/18

Reviewed by: 

Date: 4/13/18