

Datamaster Monthly Update Checklist

Department/Agency: Colchester PD DM serial #: ~~100159~~ 100157

Date: 9105 118 for 8/1 - 8/31/18

	Provided	None this month
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Breath Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Status Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Calibration	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Certification	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Installation	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Annual Preventative Maintenance	<input type="checkbox"/>	<input checked="" type="checkbox"/>
<hr/>		
Routine Performance Check	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Simulator Solution Change	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Other Reports	<input checked="" type="checkbox"/>	<input type="checkbox"/>

List other report type and date:
Diagnostic

DATAMASTER DMT OPERATOR USE LOG

Co/Chester PD
Department/Agency

100157
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
05/01/18	J. Baur	Ramon Cortez	.098 100	
05/3/18	Z. Smith	Brittany Stevens	.100	.097
5/19/18	Z. Smith	Rebecca Emery	.097	
5/24/18	Z. Smith	Wendy Mary	.098	.096
5/25/18	V. BITCA	Jessica Jensen	.098	
6/8/18	J. Brossler	Scott Provost	0.097	
06/15/18	M. West.	Michael West	.098	.096.
6/23/18	V. BITCA	Luhizo Luhizo	.097	
6/26/18	V. BITCA	Craig Hoyt	.098	
6/28/18	S. Lehmann	RPC/Solution change	.099	
7/7/18	Z. Smith	James Rand	.099	
7-15-18	Fish	Nicole O'Neill		
7-16-18	Barriere	Keith Spaulding	.099	.098
7/21/18	Z. Smith	Donna Maxwell	.098	.098
7/24/18	M. Duell	Sevren Francis	.099	.097
7/20/18	Z. Smith	Eric Fehrenbach	.097	
7/30/18	Z. Smith	Adam Desjets	.099	
8/7/18	C Jones	Dale Brooks	.097	.097
8/28/18	M. Duell	Eric Walbridge	.098	
08/29/18	Barriere	Stephen Nines	.101	

second best
direct

.097

DEVALI RESULTS LIST

09/05/2018

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
08/07/2018	22:09:11	COLCHESTER PD	CHRISTOPHER M JONES	0.187	0.169	0.097	0.097	
08/28/2018	21:58:56	COLCHESTER PD	MEGHAN J DUELL	REFUSED		0.098		
08/29/2018	03:15:32	COLCHESTER PD	JORIAN D PARRISE	0.081	0.079	0.098	0.098	
08/29/2018	23:21:34	COLCHESTER PD	JEFFREY A BAIR	0.198		0.098		RESPONSE TIMEOUT

Status Record History

Status Date	Status Time	Test Type	Status Message
08/29/2018	23:25:55	DUI Subject Record	RESPONSE TIMEOUT

DIAGNOSTIC RESULT



DataMaster DMT: 100157
Location: COLCHESTER PD
Calibration Date: 04/11/2017
Certification Date: 04/11/2017
Installation Date: 04/11/2017
Test Date: 09/01/2018
Test Time: 10:00:02

VERSIONS

DMT: 2.05
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.7°C
Breath Tube = 44.3°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.66 V
Cooler Voltage = 1.67 V
Bias Voltage = 80 V
Chopper Freq = 539 Hz

PUMP INFO

Flow Rate = 3.924 L/M

DETECTOR INFO

	ON	OFF
MAX (V)	0.0581	0.0618
MIN (V)	0.0563	0.0596

FILTER INFO

Filter 1	0.061	Zero = true
Filter 2	0.360	Zero = true
Filter 3	1.443	Zero = true

CALIBRATION CHECK

Xq = 0.093 0.09%