

Calamaster Monthly Update Checklist

Department/Agency: Colchester Police DM serial #: 100157

Date: 1/15/2019 - Dec 1 - Dec 31, 2018

	Provided	None this month
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Breath Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Status Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Calibration	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Certification	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Installation	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Annual Preventative Maintenance	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Routine Performance Check	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Simulator Solution Change	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Other Reports	<input checked="" type="checkbox"/>	<input type="checkbox"/>

List other report type and date:

Diagnostic

DATAMASTER DMT OPERATOR USE LOG

C01 Chester PD
Department/Agency

100157
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
08/29/18	J. Baur	Ryan McLeod	.198	
09/14/18	J. Baur	Craig Labeth	.127	
09/16/2018	J. Bortere	Niah Keogh	.101 .097	
9-20-18	M. Duell	Chester Cook	0.097	
9-25-18	C. Jones	David Lang	1.048	.097
10-13-18	M. Duell	Ross Pearl	.101	
10/24/18	S. GUTIERREZ	Danese Blansur	.101	.101
10/26/18	Z. Smith	Sordain Williams	.098	.097 .096
10/29/2018	J. Bortere	Wyatt Rinnig	.101 .097	.096
10/29/18	Z. Smith	Kathleen Thomas	.096	.096 ack .096
11/5/18	C. Jones	RPC/Sol Chem	1.00	
12-30-18	M. West	John Kennedy	.98	
1-5-19	Z. Smith	Jason Frye	.098	
1/09/19	K. Ray	Karen Brown	.078	

DEVALI RESULTS SUMMARY

01/15/2019

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
12/30/2018	22:38:18	COLCHESTER PD	MICHAEL S WEST	REFUSED		0.098		

Status Record History

Status Date	Status Time	Test Type	Status Message

DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location: COLCHESTER PD
Calibration Date: 04/11/2017
Certification Date: 04/11/2017
Installation Date: 04/11/2017
Test Date: 01/15/2019
Test Time: 10:00:03



VERSIONS

DMT: 2.05
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.7°C
Breath Tube = 42.8°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.66 V
Cooler Voltage = 1.67 V
Bias Voltage = 80 V
Chopper Freq = 536 Hz

PUMP INFO

Flow Rate = 4.073 L/M

DETECTOR INFO

	PUMP ON	OFF
MAX (V)	0.0421	0.0461
MIN (V)	0.0402	0.0445

FILTER INFO

Filter 1	0.045	Zero = true
Filter 2	0.342	Zero = true
Filter 3	1.426	Zero = true

CALIBRATION CHECK

Xq = 0.095 1.03%