

**Datamaster Monthly Update Checklist**

Department/Agency: Colchester Police DM serial #: 100157

Date: 8/15/18 for July 1 - July 31, 2018

	Provided	None this month
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Breath Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Status Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Calibration	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Certification	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Installation	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Annual Preventative Maintenance	<input type="checkbox"/>	<input checked="" type="checkbox"/>
<hr/>		
Routine Performance Check	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Simulator Solution Change	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Other Reports	<input type="checkbox"/>	<input type="checkbox"/>

List other report type and date:

---



---



---

**DATAMASTER DMT OPERATOR USE LOG**

Co/Chester PD  
Department/Agency

100157  
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
05/01/18	J. Baur	Ramon Cortez	.098 <del>100</del>	
05/13/18	Z. Smith	Brittany Stevens	.100	.097
5/19/18	Z. Smith	Rebecca Emery	.097 <del>100</del>	
5/24/18	Z. Smith	Wendy Mary	.098	.096
5/25/18	V. BITCA	Jessica Jensen	.098	
6/8/18	J. Bressler	Scott Probst	0.097	
06/15/18	M. West.	Michael West	.098	.096.
6/23/18	V. BITCA	Luhizo Luhizo	.097	
6/26/18	V. BITCA	Craig Hoyt	.098	
6/28/18	S. Lehmann	RPC/Solution change	.099	
7/7/18	Z. Smith	James Rand	.099	
7-15-18	<del>FISH</del>	Nicole O'Neill		
7-16-18	Barriere	Keith Spaulding	.099	.098
7/21/18	Z. Smith	Donna Maxwell	.098	.098
7/24/18	M. Duell	Sevren Francis	.099	.097
7/28/18	Z. Smith	Eric Fehrenbach	.097	
7/30/18	Z. Smith	Adam Desilets	.099	
8/7/18	C Jones	Dale Brooks	.097	.097

second best time  
.097

# Status Record History

Status Date	Status Time	Test Type	Status Message
07/21/2018	01:33:00	DUI Subject Record	Blank Error

DMT Serial Number #100157

08/15/2018

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
07/07/2018	00:51:04	COLCHESTER PD	ZACHARY M SMITH	0.156		0.099		
07/13/2018	02:59:03	COLCHESTER PD	FRANCIS C GONYAW	REFUSED		0.099		
07/15/2018	08:26:41	COLCHESTER PD	MICHAEL FISH	REFUSED		0.099		
07/15/2018	08:32:50	COLCHESTER PD	MICHAEL FISH	0.154		0.099		
07/16/2018	01:49:00	COLCHESTER PD	JORDAN D BARRIERE	0.150	0.151	0.099	0.098	
07/21/2018	01:24:20	COLCHESTER PD	ZACHARY M SMITH	0.149	0.141	0.098	0.098	Blank Error
07/24/2018	00:36:03	COLCHESTER PD	MEGHAN J DUELL	0.117	0.115	0.099	0.097	
07/28/2018	02:36:55	COLCHESTER PD	ZACHARY M SMITH	REFUSED		0.097		
07/30/2018	01:05:12	COLCHESTER PD	ZACHARY M SMITH	0.079		0.099		