

Datamaster Monthly Update Checklist

Department/Agency: Colchester PD DM serial #: 100157

Date: 4 111 12018 3-01-18 - 3-31-18

	Provided	None this month
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Breath Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Status Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Calibration	<input type="checkbox"/>	<input type="checkbox"/>
Certification	<input type="checkbox"/>	<input type="checkbox"/>
Installation	<input type="checkbox"/>	<input type="checkbox"/>
Annual Preventative Maintenance	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Routine Performance Check	<input type="checkbox"/>	<input type="checkbox"/>
Simulator Solution Change	<input type="checkbox"/>	<input type="checkbox"/>
Other Reports	<input checked="" type="checkbox"/>	<input type="checkbox"/>

List other report type and date:

Diagnostic

DATAMASTER DMT OPERATOR USE LOG

CPD

100157

Department/Agency

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
12/21/17	<i>McQueen</i>	<i>Brittany Knight</i>	<i>.99</i>	<i>99</i>
12/23/17	<i>V. Bitca</i>	<i>Brian Ploof</i>	<i>.099</i>	
12-31-17	<i>m. West</i>	<i>Keith Kennett</i>	<i>.099</i>	<i>.095</i>
01/05/18	<i>M. AKERMAN</i>	<i>STEVEN BEAUCHEMIN</i>	<i>.099</i>	
01-13-18	<i>m. West</i>	<i>Coicell Jones</i>	<i>.099</i>	
01-19-18	<i>m. West</i>	<i>David Braee</i>	<i>.098</i>	
1-19-18	<i>C Jones</i>	<i>Derick Andors</i>	<i>.098</i>	<i>.097</i>
2-5-18	<i>Z Smith</i>	<i>Jonathan Pratt</i>	<i>.098</i>	<i>.099</i>
2-5-18	<i>Z Smith</i>	<i>Jonathan Pratt</i>	<i>.099</i>	<i>.098</i>
2-11-18	<i>Z Smith</i>	<i>Matthew Line</i>	<i>.098</i>	<i>.097</i>
2-11-18	<i>Z Smith</i>	<i>Matthew Line</i>	<i>.097</i>	
02-11-18	<i>m. West</i>	<i>Peter Laddasack</i>	<i>.098</i>	<i>.096</i>
2-28-18	<i>C Jones</i>	<i>RPC Sol, change</i>	<i>.100</i>	
04/02/18	<i>J. Baer</i>	<i>Richard Cormier</i>	<i>.100⁹⁹</i>	
4-8-18	<i>V. Bitca</i>	<i>Shane Cannon</i>	<i>.098</i>	
4/9/18	<i>R. Dorell</i>	<i>TEST/APM</i>	<i>.099</i>	<i>—</i>

DATAMASTER DMT MAINTENANCE LOG

Calchester PD
Department/Agency

100157
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/31/17	<p style="text-align: center;"><i>Installed version 2.00 RPC chip.</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p style="text-align: center;"><i>R. Driscoll</i> VFL</p>
2-28-18	<p style="text-align: center;">DMT Simulator Solution Lot # 17-18-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 05/02/2017 Expires: 05/02/2018</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">OPENED: <i>2-28-18</i></p> <p style="text-align: center;"><i>RPC Solution Change</i></p>	<p style="text-align: center;"><i>C Jones</i> 181</p>
4/9/18	<p style="text-align: center;"><i>Completed APMT Protocol.</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p style="text-align: center;"><i>R. Driscoll</i> VFL</p>
	<p>[For RPC or Simulator Solution Change Place Label Here]</p>	

DIAGNOSTIC RESULTS

04/11/2018

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
03/25/2018	19:35:41	COLCHESTER PD	FRANCIS C GONYAW	0.128	0.129	0.099	0.099	

Status Record History

Status Date	Status Time	Test Type	Status Message

ANNUAL PREVENTATIVE MAINTENANCE



DataMaster DMT: 100157
Location: COLCHESTER PD
Calibration Date: 04/11/2017
Certification Date: 04/11/2017
Installation Date: 04/11/2017
APM Date: 04/09/2018
Performed by: ROBERT DRISCOLL

Diagnostic Results

VERSIONS
DMT: 2.03
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 49.4°C
Breath Tube = 42.8°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.66 V
Cooler Voltage = 1.67 V
Bias Voltage = 80 V
Chopper Freq = 533 Hz

PUMP INFO

Flow Rate = 4.059 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX (V)	-0.0024	0.0018
MIN (V)	-0.0041	0.0000

FILTER INFO

Filter	Value	Zero
Filter 1	0.001	Zero = true
Filter 2	0.300	Zero = true
Filter 3	1.396	Zero = true

CALIBRATION CHECK

Xq = 0.093 0.37%

APM PASSED

Simulator Vapor Test


Concentration = 0.100 g/210L
Lot # = 17-18-100
Range = 0.095 - 0.105
Simulator Vapor = 0.099 g/210L

RF Detection Test

Passed

Sample Acceptance Test

Passed



Performed by _____

Date _____

04/09/2018

Reviewed by _____

Date _____

DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location: COLCHESTER PD
Calibration Date: 04/11/2017
Certification Date: 04/11/2017
Installation Date: 04/11/2017
Test Date: 04/01/2018
Test Time: 10:00:02



VERSIONS

DMT: 2.03
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.7°C
Breath Tube = 44.6°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.66 V
Cooler Voltage = 1.67 V
Bias Voltage = 80 V
Chopper Freq = 542 Hz

PUMP INFO

Flow Rate = 3.989 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX(V)	0.0134	0.0177
MIN(V)	0.0118	0.0150

FILTER INFO

Filter 1	0.016	Zero = true
Filter 2	0.315	Zero = true
Filter 3	1.414	Zero = true

CALIBRATION CHECK

Xq = 0.094 0.37%