

Datamaster Monthly Update Checklist

Department/Agency: Colchester Police DM serial #: 100157

Date: 6111 118 - May 1 - May 31, 2018

	Provided	None this month
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Breath Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Status Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Calibration	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Certification	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Installation	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Annual Preventative Maintenance	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Routine Performance Check	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Simulator Solution Change	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Other Reports	<input type="checkbox"/>	<input checked="" type="checkbox"/>

List other report type and date:

DELAWARE INDUSTRIES

06/11/2018

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
05/01/2018	08:30:11	COLCHESTER PD	JEFFREY A BAUR	0.130		0.098		
05/03/2018	00:36:21	COLCHESTER PD	ZACHARY M SMITH	0.201	0.186	0.099	0.097	
05/19/2018	03:22:17	COLCHESTER PD	ZACHARY M SMITH	0.102		0.097		
05/24/2018	13:42:54	COLCHESTER PD	ZACHARY M SMITH	0.268		0.098	0.096	RESPONSE TIMEOUT
05/24/2018	13:53:33	COLCHESTER PD	ZACHARY M SMITH	0.250		0.097		Blank Error
05/25/2018	00:05:07	COLCHESTER PD	VICTOR BITCA	0.123		0.098		

Status Record History

Status Date	Status Time	Test Type	Status Message
05/24/2018	13:49:58	DUI Subject Record	RESPONSE TIMEOUT
05/24/2018	13:57:13	DUI Subject Record	Blank Error