

Calamaster Monthly Update Checklist

Department/Agency Colchester Police Dept DM serial #: 100157

Date: 12 104 118 for Nov 1-30, 2018

	Provided	None this month
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Breath Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Status Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Calibration	<input type="checkbox"/>	<input type="checkbox"/>
Certification	<input type="checkbox"/>	<input type="checkbox"/>
Installation	<input type="checkbox"/>	<input type="checkbox"/>
Annual Preventative Maintenance	<input type="checkbox"/>	<input type="checkbox"/>
Routine Performance Check	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Simulator Solution Change	<input type="checkbox"/>	<input type="checkbox"/>
Other Reports	<input type="checkbox"/>	<input type="checkbox"/>

List other report type and date:

Diagnostic

DMT Serial Number #100157

12/04/2018

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message

Status Record History

Status Date	Status Time	Test Type	Status Message

DATAMASTER DMT OPERATOR USE LOG

Co/Chester PD
Department/Agency

100157
DataMaster DMT Serial Numt

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
08/29/18	J. Baur	Ryan McLeod	.198	
09/14/18	J. Baur	Craig Labeth	.127	
09/16/2018	J. Bortere	Niah Keogh	.107 .097	
9-20-18	M. Duell	Chester Cook	0.097	
9-25-18	C. Jones	David Lang	1.048	.047
10-13-18	M. Duell	Ross Pearl	.101	
10/24/18	S. GUTIERREZ	Daniel Blansur	.101	.101
10/26/18	Z. Smith	Jordain Williams	.098	.097 .097
10/29/2018	J. Bortere	Wyatt Ranning	.107 .097	.096
10/29/18	Z. Smith	Kathleen Thomas	.096	.096 .096
11/5/18	C. Jones	RPC/SOI Chem	1.100	

DATAMASTER DMT MAINTENANCE LOG

Colchester PD
Department/Agency

100157
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/31/17	<p style="text-align: center;"><i>Installed version 2.08 RPC chip.</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>R. Driscoll</i> VFL</p>
2-28-18	<p style="text-align: center;">DMT Simulator Solution Lot # 17-18-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 05/02/2017 Expires: 05/02/2018</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">OPENED: <i>2-28-18</i></p> <p style="text-align: center;"><i>RPC Solution Change</i></p>	<p><i>C Jones</i> 181</p>
4/9/18	<p style="text-align: center;"><i>Completed APM Protocol.</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>R. Driscoll</i> VFL</p>
6/28/18	<p style="text-align: center;">DMT External Std Solution Lot: 18-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/28/18 Exp Date: 02/28/19 Opened: <i>6/28/18</i></p> <p style="text-align: center;"><i>RPC Solution Change</i></p>	<p><i>J. Lehmann</i> M193</p>
11-5-18	<p style="text-align: center;">DMT External Std Solution Lot: 18-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/28/18 Exp Date: 02/28/19 Opened: <i>11-5-18</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;"><i>RPC / Solution Change</i></p>	<p><i>C Jones</i> M181</p>

ROUTINE PERFORMANCE CHECK REPORT



DataMaster DMT: 100157
Location: COLCHESTER PD
Calibration Date: 04/11/2017
Certification Date: 04/11/2017
Installation Date: 04/11/2017
RPC Date: 11/05/2018
Supervisor Name: CHRISTOPHER M JONES

Diagnostic Results

VERSIONS
DMT: 2.05
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.9°C
Breath Tube = 46.0°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.66 V
Cooler Voltage = 1.67 V
Bias Voltage = 80 V
Chopper Freq = 543 Hz

PUMP INFO

Flow Rate = 3.966 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX (V)	0.0736	0.0768
MIN (V)	0.0720	0.0752

FILTER INFO

Filter 1	0.076	Zero = true
Filter 2	0.373	Zero = true
Filter 3	1.462	Zero = true

CALIBRATION CHECK

Xq = 0.093 0.28%

Routine Performance Check Passed

Accuracy and Precision Check


Concentration	= 0.101 g/210L
Lot #	= 18-04-100
Range	= 0.096 - 0.106
Average	= 0.099 g/210L
Std Dev	= 0.0004

RF Detection Test

Passed

Sample Acceptance Test

Passed



Performed by Chris Jones 181

Date 11/05/2018

Reviewed by _____

Date _____

DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location: COLCHESTER PD
Calibration Date: 04/11/2017
Certification Date: 04/11/2017
Installation Date: 04/11/2017
Test Date: 12/01/2018
Test Time: 10:00:03



VERSIONS

DMT: 2.05
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.7°C
Breath Tube = 44.1°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.66 V
Cooler Voltage = 1.67 V
Bias Voltage = 80 V
Chopper Freq = 543 Hz

PUMP INFO

Flow Rate = 4.062 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX(V)	0.0665	0.0703
MIN(V)	0.0652	0.0685

FILTER INFO

Filter 1	0.069	Zero = true
Filter 2	0.364	Zero = true
Filter 3	1.454	Zero = true

CALIBRATION CHECK

Xq = 0.094 0.37%