

Datamaster Monthly Update Checklist

Department/Agency: Colchester Police DM serial #: 100157

Date: 11/02/2018 for 10/1/18 - 10/31/18

	Provided	None this month
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Breath Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Status Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Calibration	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Certification	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Installation	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Annual Preventative Maintenance	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Routine Performance Check	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Simulator Solution Change	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Other Reports	<input checked="" type="checkbox"/>	<input type="checkbox"/>

List other report type and date:

Diagnostic

DL26411 REVULU 1125017

11/02/2018

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
10/13/2018	00:16:55	COICHESTER PD	MEGHAN J DUELL	0.167		0.098		
10/24/2018	01:15:38	COICHESTER PD	STEPHEN J GUTIERREZ	0.100	0.101	0.097	0.096	
10/26/2018	03:19:16	COICHESTER PD	ZACHARY M SMITH	0.191		0.098	0.095	Simulator Out of Range.
10/26/2018	03:27:10	COICHESTER PD	ZACHARY M SMITH	0.192	0.187	0.097	0.096	
10/29/2018	16:56:55	COICHESTER PD	JORDAN D BARRIERE	0.122	0.115	0.097	0.096	
10/29/2018	20:11:02	COICHESTER PD	ZACHARY M SMITH	0.134		0.096	0.094	
10/29/2018	20:20:06	COICHESTER PD	ZACHARY M SMITH	0.136		0.096		Simulator Out of Range

DMT Serial Number #100157

Status Record History

Status Date	Status Time	Test Type	Status Message
10/26/2018	03:23:51	DUI Subject Record	Simulator Out of Range
10/29/2018	20:16:46	DUI Subject Record	Simulator Out of Range

DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location: COLCHESTER PD
Calibration Date: 04/11/2017
Certification Date: 04/11/2017
Installation Date: 04/11/2017
Test Date: 10/15/2018
Test Time: 10:00:02



VERSIONS

DMT: 2.05
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.7°C
Breath Tube = 46.8°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.66 V
Cooler Voltage = 1.67 V
Bias Voltage = 80 V
Chopper Freq = 543 Hz

PUMP INFO

Flow Rate = 3.862 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX (V)	0.0797	0.0823
MIN (V)	0.0772	0.0803

FILTER INFO

Filter 1	0.081	Zero = true
Filter 2	0.379	Zero = true
Filter 3	1.467	Zero = true

CALIBRATION CHECK

Xq = 0.094 0.56%