



Simulator Temperature Check

Simulator Serial Number:

MP3508

DMT Serial Number / Simulator Location:

100157 / Colchester PD

NIST Traceable Thermometer Serial Number:

160294076

As Found

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator	33.95	34.00	0.05	Y

\* 34.0 +/- 0.2 degree Celsius acceptable range

Was temperature adjustment performed?

Y / (N)

As Left (if applicable)

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator				

Notes:

Check Performed By:

 JSO

Date:

3/11/19

Reviewed by:



Date:

3/13/19