

Datamaster Monthly Update Checklist

Department/Agency: Colchester PD DM serial #: 100157

Date: 11/07/19 for 10-01-19 to 10-31-19

	Provided	None this month
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Breath Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Status Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Calibration	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Certification	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Installation	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Annual Preventative Maintenance	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Routine Performance Check	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Simulator Solution Change	<input type="checkbox"/>	<input type="checkbox"/>
Other Reports	<input checked="" type="checkbox"/>	<input type="checkbox"/>

List other report type and date:

Diagnostic Result

DATAMASTER DMT OPERATOR USE LOG

Colchester Police

Department/Agency

100157

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
6-26-19	Chris Jones	RPC/Solcham		
06-28-19	czachor	Hanna Woodard	1100	
06-29-19	czachor	Amanda Bae	1099	
07/01/19	czachor	Bleonda Sulejmani	1093 1099	
07/28/19	Bitca, Victor	Jennifer Macdonald	100	
07/20/19	czachor	Andrew McDonald	1103	
8-4-19	Duell	David Saffield	098	098
8-18-19	Fell	Mark Vuley	098	098, 097
10-8-19	CHANG BPP	Megan Magnus	098 1093	
10/18/19	Fell	Nathan Shaffer	097	
10/17/19	GUTIERREZ	CINDY BERNICE	098	
10-19-19	mwest	RPC/Solomia chag		
10/19/19	BITCA	Sven Morway	092	
10/27/19	GUTIERREZ	RICHARDSON, K	099	
10/29/19	Fell	Mika, S	099	
11/1/19	GUTIERREZ	FISHER, H.	098	

Colchester Police

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

100157

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/28/19	<p align="center">DMT External Std Solution</p> <p>Lot: 18-05-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/21/18 Exp Date: 03/21/19 Opened: 2-28-19 VB</p> <p>[For RPC Range: 0.095 - 0.105 e Label Here]</p> <p align="center">RPC / Solution Change</p>	V. Bitca M194 CPD
3/11/19	<p align="center">DMT External Std Solution</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	J Dukette CVFL
6-26-19	<p align="center">DMT External Std Solution</p> <p>Lot: 19-07-100 EtOH Certified Conc. 0.103 Range: 0.098 - 0.108 Prep Date: 5/31/19 Exp Date: 05/31/20 Opened: 6-26-19</p> <p>[For R Range: 0.098 - 0.108 lace Label Here]</p> <p align="center">RPC Solution Change</p>	M181 C Jones
10-19-19	<p align="center">DMT External Std Solution</p> <p>Lot: 19-07-100 EtOH Certified Conc. 0.103 Range: 0.098 - 0.108 Prep Date: 5/31/19 Exp Date: 05/31/20 Opened: 10-19-19 MSW</p> <p>[FOI Range: 0.098 - 0.108 e Place Label Here]</p> <p align="center">RPC / Solution Change</p>	M192 M West

DETAILED RESULTS HISTORY

11/07/2019

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
10/08/2019	04:26:29	BURLINGTON PD	CALEN CHANG	0.199		0.097		Simulator Out of Range
10/08/2019	04:32:19	BURLINGTON PD	CALEN CHANG	REFUSED		0.098		
10/17/2019	21:32:31	COLCHESTER PD	STEPHEN J GUILHERREZ	REFUSED		0.098		
10/18/2019	22:44:55	COLCHESTER PD	ELIZABETH M FELL			0.097		Simulator Out of Range
10/18/2019	23:53:21	COLCHESTER PD	ELIZABETH M FELL			0.097		Simulator Out of Range
10/19/2019	15:12:58	COLCHESTER PD	VICTOR BITCA			0.097		Simulator Out of Range
10/27/2019	05:35:06	COLCHESTER PD	STEPHEN J GUILHERREZ	REFUSED		0.099		
10/29/2019	21:50:35	COLCHESTER PD	ELIZABETH FELL	REFUSED		0.099		

DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location: COLCHESTER PD
Calibration Date: 04/11/2017
Certification Date: 04/11/2017
Installation Date: 04/11/2017
Test Date: 11/01/2019
Test Time: 10:00:02



VERSIONS

DMT: 2.05
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 49.0°C
Breath Tube = 44.6°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.66 V
Cooler Voltage = 1.67 V
Bias Voltage = 80 V
Chopper Freq = 541 Hz

PUMP INFO

Flow Rate = 4.016 L/M

DETECTOR INFO

	ON	OFF
MAX(V)	0.1285	0.1325
MIN(V)	0.1262	0.1305

FILTER INFO

Filter 1	0.132	Zero = true
Filter 2	0.427	Zero = true
Filter 3	1.496	Zero = true

CALIBRATION CHECK

Xq = 0.095 1.50%