

Datamaster Monthly Update Checklist

Department/Agency: Colchester Police DM serial #: 100157

Date: 5 101 12020 4/01/20 - 4/30/20

	Provided	None this month
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Breath Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Status Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Calibration	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Certification	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Installation	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Annual Preventative Maintenance	<input type="checkbox"/>	<input checked="" type="checkbox"/>
<hr/>		
Routine Performance Check	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Simulator Solution Change	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Other Reports	<input type="checkbox"/>	<input type="checkbox"/>

List other report type and date:

DATAMASTER DMT OPERATOR USE LOG

Colchester PD
Department/Agency

100157
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
12/29	czachor	Booska	.1098	
1-16-20	Duell	Delucco	.098	
1-16-2020	West	K. Hell	.098	
1-18-2020	Duell	Bugbee	.098	.098
1/26	Giroux	Reyome	.098	
2/14/2020	S. Stratton (VFL)	APM/Test	0.097	
2/14/2020	S. Stratton (VFL)	APM/Test	0.098	
2-22-20	Fish	Myers		
2-22-2020	Jones	RPC / sulphur	.1102	
3/19/20	Smith	Delgado	.099	
3-31-20	Duell	Green	.096	
4-4-20	Duell	Griffin	.097	.096
4-15-20	K. Sylvia CCSD	Larabee	.099	
4/25/20	Felt	Andrews	.098	.098
4/26/20	Smith	Rotunno	.98	.097

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
04/04/2020	02:11:44	COLCHESTER PD	MEGHAN J DUELL	0.137		0.097	0.094	Simulator Out of Range
04/04/2020	02:19:09	COLCHESTER PD	MEGHAN J DUELL	0.134		0.096		
04/15/2020	18:16:58	CHITTENDEN SHERIFF	ROBERT SYLVIA	0.170		0.099		
04/25/2020	22:139:13	COLCHESTER PD	E FELL	INVALID		0.098		
04/25/2020	22:159:52	COLCHESTER PD	E FELL	0.105		0.098		
04/26/2020	23:15:27	COLCHESTER PD	ZACHARY M SMITH	0.138	0.141	0.098	0.097	
04/29/2020	03:17:17	COLCHESTER PD	D C DEWEY	REFUSED		0.098		

