

**Datamaster Monthly Update Checklist**

Department/Agency: Colchester PD

DM serial #: 100157

Date: 7 / 10 / 2020 June 1 - June 30, 2020

	Provided	None this month
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Breath Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Status Summary Report	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Calibration	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Certification	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Installation	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Annual Preventative Maintenance	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Routine Performance Check	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Simulator Solution Change	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Other Reports	<input type="checkbox"/>	<input checked="" type="checkbox"/>

List other report type and date:

---



---



---

**DATAMASTER DMT OPERATOR USE LOG**

Colchester PD

Department/Agency

100157

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
12/29	czachor	Booska	.098	
1-16-20	Duell	Delucco	.098	
1-16-2020	west	K. Htell	.098	
1-18-2020	Duell	Bugbee	.098	.098
1/26	Giroux	Reyome	.098	
2/14/2020	S. Stratton (VFL)	APM / Test	0.097	
2/14/2020	S. Stratton (VFL)	APM / Test	0.098	
2-22-20	Fish	Miyers		
2-22-2020	Jones	RPC / solphunx	.100	
3/19/20	Smith	Delgado	.098	
3-31-20	Duell	Green	.096	
4-4-20	Duell	Giffin	.097	.096
4-15-20	R. Sylvia CCSD	Larabee	.098	
4/25/20	Felt	Andrews	.098	.098
4/26/20	Smith	Rotunno	.98	.097
5/22/20	Gutierrez, S.	EDELKAMP, K.	.100	
5/25/20	AKERMAN, M.	COME, JAMES D.	.097	.097
05/30/20	Ray, Z	ELSTIS, JESSICA F.	.097	.097
6/4/20	Giroux	Gonzales	.100	
6/23/20	Smith	Abdikadir	.096	



# DATAMASTER DMT MAINTENANCE LOG

Colchester PD

Department/Agency

100157

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/14/2020	<p>Completed APM protocol                      [ For RPC or Simulator Solution Change Place Label Here]                      - 1<sup>st</sup> attempt sim out of range                      - 2<sup>nd</sup> attempt passed</p>	S. Stratton (VFL)
<b>DMT External Std Solution</b>		
2-21-2020	<p>Lot: 19-13-100 EtOH Certified Conc. 0.100                      Range: 0.095 - 0.105                      Prep Date: 9/23/19 Exp Date: 09/23/20                      [ For RPC Opened: 2-21-2020 cm] ace Label Here]                      Solution change / RPC</p>	C Jones 181
<b>DMT External Std Solution</b>		
7-3-2020	<p>[ For RPC Lot: 20-03-100 EtOH Certified Conc. 0.100                      Range: 0.095 - 0.105                      Prep Date: 1/29/20 Exp Date: 01/29/21                      Opened: 7-3-2020 cm] e Label Here]                      Solution Change / RPC</p>	C Jones 181
	[ For RPC or Simulator Solution Change Place Label Here]	

# Status Record History

Status Date	Status Time	Test Type	Status Message

DELAWARE DELIVERY

07/09/2020

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
06/04/2020	16:01:46	COLCHESTER PD	HIDELFO GONZALES	0.120	0.111	0.096	0.095	
06/23/2020	02:21:24	COLCHESTER PD	ZACHARY M SMITH	REFUSED		0.096		