



# SOLUTION CHANGE

DataMaster DMT: 100157  
Location: COLCHESTER PD  
Date: 11/02/2021  
Performed by: CHRISTOPHER M JONES



## Accuracy and Precision Check

Concentration = 0.100 g/210L  
Lot # = 25-15-100  
Range = 0.095 - 0.105  
Average = 0.099 g/210L  
Std Dev = 0.0005

Simulator Temperature: 34.0°C

\* incorrect lot.  
SSC not needed.  
RPC performed w/  
passing results. No tests  
run b/w this SSC  
and passing RPC  
JSD 11/2/21

Performed by

*Christopher M Jones*

Date

11/02/2021

**State of Vermont**  
**DMHost for Windows**

Breath Test Summary

SerialNumber = '100157' and TestDate = '11/02/2021'

Serial #	Test Date/Time	Sample 1	Sample 2	Sample 3	Sample 4	Sample 5	Sample 6	Sim 1	Sim 2	Sim T*	Sim Lot	Target Status Code
----------	----------------	----------	----------	----------	----------	----------	----------	-------	-------	--------	---------	--------------------

End of Report

100157

DFP