

ROUTINE PERFORMANCE CHECK REPORT



DataMaster DMT: 100158
Location: Northfield PD
Calibration Date: 11/16/2011
Certification Date: 11/17/2011
Installation Date: 12/08/2011
RPC Date: 02/16/2014
Supervisor Name: CHRISTOPHER M QUESNEL

Diagnostic Results

VERSIONS

DMT: 1.01
PIC: 2.05
Modem: 2.4
Questions: 2.1

TEMPERATURES

Sample Chamber = 49.1°C
Breath Tube = 43.3°C
Digital Sim = 34.1°C

SETTINGS

Lamp Voltage = 1.76 V
Cooler Voltage = 1.70 V
Bias Voltage = 80 V
Chopper Freq = 550 Hz

PUMP INFO

Flow Rate = 4.989 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX (V)	0.0787	0.0825
MIN (V)	0.0764	0.0803

FILTER INFO

Filter 1	0.082	Zero = true
Filter 2	0.380	Zero = true
Filter 3	1.376	Zero = true

CALIBRATION CHECK

Xq = 0.092 2.42%

Accuracy and Precision Check


Concentration	= 0.099 g/210L
Lot #	= 13-16-100
Range	= 0.094 - 0.104
Average	= 0.096 g/210L
Std Dev	= 0.0000

RF Detection Test

Passed

Sample Acceptance Test

Passed



Routine Performance Check Passed

Performed by

Ch. Quesnel

Date

02/16/2014

Reviewed by

[Signature]

Date

2/19/14

DMT Serial Number #100158

Page 1 of 1

02/16/2014 4:57 PM

[Signature]
2/19/2014