

Status Record History

Status Date	Status Time	Test Type	Status Message

Log On Record History

Test Date	Test Time	Name

DATAMASTER DMT MAINTENANCE LOG

Berlin PD
Department/Agency

100148
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/12/14	<p>Completed APT Protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p>R. Driscoll DMT TECH.</p>
2/7/15	<p>DataMaster Simulator Solution</p> <p>Lot # 14-10-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 7/14/2014 Expires: 7/14/2015</p> <p>OPENED: <u>CS 2/7/15</u></p> <p style="font-size: 2em; margin-top: 20px;">RPC</p>	
3.9.15	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="font-size: 1.5em; margin-top: 20px;">Changed Clock + 9 min</p>	<p>MM</p>
6/13/15	<p>DataMaster Simulator Solution</p> <p>Lot # 15-01-100 EtOHaq Conc: 0.098 g/210L Range: 0.093 - 0.103 g/210L Preparation Date: 1/21/2015 Expires: 1/21/2016</p> <p>OPENED: <u>6/13/15 CS</u></p> <p style="font-size: 2em; margin-top: 20px;">RPC</p>	

Sellin

DATAMASTER DMT OPERATOR USE LOG

Department/Agency

100148
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
8/30/14	Bullard, Jonathan	Andrew Barr	.100	.099
9/15/14	RPC Bassett, Chad	Tyle/Eckstein	.096	
9-22-14	Kapitaniski, Kyle	Justin Galois	.098	
10/25/14	RPC	RPC	.099	
10/27/14	Quesnel, Chris	Robert Richardson	.097	
11/2/14	Bullard, Jonathan	James Thompson	.097	
11/12/14	R. Dascoll	TEST / ARM	.098	-
11/15/14	Bullard, Jonathan	Thomas Warden	.097	.096
11/22/14	Bullard, Jonathan	Adrianna Milo	.097	.096
12-19-14	Montelth, Mark	Robert Amburn	.095	.095
2/7/14	Bassett, Chad	RPC	.098	
4-27-15	Friedholm, Jon	Training	.097	
6/13/15	Bassett, Chad	RPC	.099	