



**Northfield**  
Police Department

51 South Main Street  
Northfield, VT 05663  
(802) 485-9181  
FAX (802) 485-4188

## DATAMASTER DMT Monthly Update Checklist

DATAMASTER SERIAL # 100158

MONTH: JUNE / 2015

	PROVIDED	NONE THIS MONTH
Operators Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Maintenance Logbook	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Breath Summary Report	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Status Summary Report	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Calibration Report	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Certification Report	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Installation Report	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Annual Preventative Maintenance	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Routine Performance Check	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Simulator Solution Change	<input type="checkbox"/>	<input checked="" type="checkbox"/>

Other Reports

List other report type and date:

---



---



---



---



---

DATE FAXED TO WCSA: \_\_\_\_ / \_\_\_\_ / \_\_\_\_

*"With Honor, We Serve"*

**DATAMASTER DMT OPERATOR USE LOG**

Northfield PD  
Department/Agency

100158  
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
8/8	Linda Shedd	Canas, Michael	.097	
8/8	Linda Shedd	Canas, Michael	.097	.096
9-17	Linda Shedd	Flaherty, Matthew	.095	
9/19/14	Quesnel, Christopher M.	Sim Solution Change	.097	
10/27/14	Carriveau, Joseph	Cruidshank, Edwin	.097	
10/30/14	Quesnel, Christopher M.	October R.P.C.	.097	
11/12/14	R. Driscoll	TEST /APM	.096	—
11/23/14	Leary, Nevin S. ←	→ Pickel, J.	.096	.094
12/2/14	Carriveau, Joseph <small>KTucker &amp;</small>	Hartman, Matthew	.097	
12/08/14	Carriveau, Joseph	Potter, Justin	.095	
12/21/14	Withrow, Daniel	Bedell, Amanda	OUT OF RANGE	
12/26/14	Quesnel, Christopher M.	Sim Solution Change	.097	—
1/20/15	Hear, Brian	James H. Morsha	.095	
2/12/15	Hear, Brian	James Smith Jr.	.097	
2/12/15	Hear, Brian	James Smith Jr.	.097	
2/22/15	Quesnel, Christopher, M.	Feb. R.P.C.	.097	
3-7-15	Pouliot, A	Gilbert, Jonathan	.096	.095
4/19/15	Withrow, Daniel	Fletcher, Harold	.096	
6/27/15	Quesnel, Christopher M.	June R.P.C.	.098	
7/9/15	Withrow, Daniel	Tolan, C, Jameson	.098	

# DATAMASTER DMT MAINTENANCE LOG

Northfield PD

100158

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
12/26/14	<p><b>DataMaster Simulator Solution</b></p> <p>[ For RPC or ] Lot # 14-10-100 EtOHaq Conc: 0.097 Label Here]</p> <p>Range: 0.092 - 0.102</p> <p>Preparation Date: 7/14/2014 Expires: 7/14/2015</p> <p>OPENED: 12/26/14 <span style="border: 1px solid black; border-radius: 50%; padding: 2px;">CMB</span></p> <p style="font-size: 1.5em; font-family: cursive; text-align: center;">SIM. SOLUTION CHANGE</p>	Chris  Desmet
2/22/15	<p><b>DataMaster Simulator Solution</b></p> <p>[ For RPC or ] Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Label Here]</p> <p>Range: 0.094 - 0.104</p> <p>Preparation Date: 12/9/2014 Expires: 12/9/2015</p> <p>OPENED: 2/22/15 <span style="border: 1px solid black; border-radius: 50%; padding: 2px;">CMB</span></p> <p style="font-size: 1.5em; font-family: cursive; text-align: center;">FEB. 2015 R.P.C.</p>	Chris  Desmet
6/27/15	<p><b>DataMaster Simulator Solution</b></p> <p>[ For RPC or ] Lot # 15-07-100 EtOHaq Conc: 0.100g/210L Label Here]</p> <p>Range: 0.095 - 0.105</p> <p>Preparation Date: 5/11/2015 Expires: 5/11/2016</p> <p>OPENED: 6/27/15 <span style="border: 1px solid black; border-radius: 50%; padding: 2px;">CMB</span></p> <p style="font-size: 1.5em; font-family: cursive; text-align: center;">JUNE 2015 R.P.C.</p>	Chris  Desmet
	[ For RPC or Simulator Solution Change Place Label Here]	

