

SOLUTION CHANGE

DataMaster DMT: 100158
Location: Northfield PD
Date: 02/20/2016
Performed by: MICHAEL R GERO



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-13-100
Range = 0.094 - 0.104
Average = 0.098 g/210L
Std Dev = 0.0000

RCD
2/22/16

Simulator Temperature: 34.0°C

Performed by *Michael R Gero*

Date 02/20/2016