

SOLUTION CHANGE

DataMaster DMT: 100158
Location: NORTHFIELD PD
Date: 04/29/2020
Performed by: MICHAEL R GERO



Accuracy and Precision Check

Concentration = 0.101 g/210L
Lot # = 20-04-100
Range = 0.096 - 0.106
Average = 0.098 g/210L
Std Dev = 0.0004

SJS 05/01/2020

Simulator Temperature: 34.0°C

Performed by

Michael R Gero

Date

04/29/2020

SOLUTION CHANGE

DataMaster DMT: 100158
Location: NORTHFIELD PD
Date: 04/25/2020
Performed by: MICHAEL R GERO



Accuracy and Precision Check

Concentration = 0.101 g/210L
Lot # = 18-11-100
Range = 0.096 - 0.106
Average = 0.097 g/210L
Std Dev = 0.0004

Simulator Temperature: 34.0°C

*Expired sol'n.
Called to have
instrument removed
from service 4/27/2020 JSO*

Performed by

Michael R Gero

Date

04/25/2020

**State of Vermont
DMHost for Windows**

Breath Test Summary

SerialNumber = '100158' and TestDate between '4/25/2020' and '4/29/2020'

Serial #	Test Date/Time	Sample 1	Sample 2	Sample 3	Sample 4	Sample 5	Sample 6	Sim 1	Sim T*	Sim 2	Sim T*	Sim Lot	Target Status Code
100158	04/25/2020 00:57							0.098	34.0			19-07-100	0.103 103
100158	04/25/2020 01:21							0.096	34.0			19-07-100	0.103 11
100158	04/25/2020 01:25							0.096	34.0			19-07-100	0.103 11
100158	04/25/2020 01:28							0.097	34.0			19-07-100	0.103 11
100158	04/25/2020 01:33							0.097	34.0			19-07-100	0.103 11
100158	04/25/2020 01:38							0.097	34.0			19-07-100	0.103 11
100158	04/25/2020 01:43							0.097	34.0			19-07-100	0.103 11

End of Report

100158 *SSB* *Breath tests all completed prior to SSC with expired solution. No breath tests prior to passing SSC. SSB 05/01/2020*

Re: DMT Mandatory Training Update

DPS - DMT <DPS.DMT@vermont.gov>

Mon 4/27/2020 2:54 PM

To: Gero, Michael <Michael.Gero@vermont.gov>

Afternoon,

I was reviewing the records this morning and noticed you performed the solution change using 18-11-100. This is an expired lot of sim solution. I contacted Sgt Hoar this morning to request a note be placed on the DMT. The DMT will need to be removed from service until the solution can be replaced. Sim solution was sent out this morning. You should have it tomorrow or Wednesday. You can also borrow some from a different agency if you'd like it to get back online before then. Berlin or Barre City should have some.

Get [Outlook for iOS](#)

From: DPS - DMT <DPS.DMT@vermont.gov>**Sent:** Monday, April 27, 2020 6:57:50 AM**To:** Gero, Michael <Michael.Gero@vermont.gov>**Subject:** Re: DMT Mandatory Training Update

I will have those sent out today.

Thanks,

Jeff

From: Gero, Michael <Michael.Gero@vermont.gov>**Sent:** Saturday, April 25, 2020 7:21 PM**To:** DPS - DMT <DPS.DMT@vermont.gov>**Subject:** Re: DMT Mandatory Training Update

Hello;

May I please get a couple of Simulator Solution Bottles ASAP? My DMT went out of range as Lots 19-07-100 and 18-11-100 tested out at the low side of the range spread: .097 within a .096-.106 range.

Get [Outlook for iOS](#)

From: DPS - DMT <DPS.DMT@vermont.gov>**Sent:** Monday, April 13, 2020 6:05:03 PM**To:** Trudo, Brittany <Brittany.Trudo@vermont.gov>; Eastman, Larry <Larry.Eastman@vermont.gov>; Pickel, Justin <Justin.Pickel@vermont.gov>; Pierce, Joel <Joel.Pierce@vermont.gov>; Kirkpatrick, Troy <Troy.Kirkpatrick@vermont.gov>; Murphy, Brianna <Brianna.Murphy@vermont.gov>; Keefe, Michael <Michael.Keefe@vermont.gov>; Tuttle, Ian <Ian.Tuttle@vermont.gov>; Parro, Joshua <Joshua.Parro@vermont.gov>; Joel.Howard@benningtonsheriff.org <Joel.Howard@benningtonsheriff.org>; Stemp, Joshua <Joshua.Stemp@vermont.gov>; Lauzon, Chris <Chris.Lauzon@vermont.gov>; Tiersch, Steven <Steven.Tiersch@vermont.gov>; Bassette, Chad <Chad.Bassette@vermont.gov>; Monteith, Mark