

Simulator Temperature Check

Simulator Serial Number: MP3455

DMT Serial Number / Simulator Location: 100158 / Northfield PD

NIST Traceable Thermometer Serial Number: Z1FMM02249

As Found

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator	33.99	33.99	0.00	Y

* $\pm 0.10^{\circ}\text{C}$ agreement between thermometer and simulator
 * $34.0 \pm 0.2^{\circ}\text{C}$ acceptable range

Was temperature adjustment performed? Y / (N)

As Left (if applicable)

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator				

Notes:

Check Performed By: [Signature]

Reviewed by: [Signature]

Date: 7/8/22

Date: ~~7/8/22~~ 11/2/22

[Signature]