

DATAMASTER DMT MAINTENANCE LOG

12/20/11 - 1/24/12

Milton Pol

100159

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>12/20/11</i>	DataMaster Simulator Solution Lot # 12-38-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 11/17/2011 Expires: 11/17/2012 []nge Place Label Here] OPENED: <i>12/20/11</i> <i>Install Per Protocol</i> <i>Sam 12/20/11</i>	<i>Steven Harnois</i>
<i>3 Feb 12</i>	<i>RPC - Passed</i> [For RPC or Simulator Solution Change Place Label Here] <i>After RPC DataMaster restart "Due to Fatal error" TWICE. When attempting to print 0.012 reports</i>	<i>Tom DePant</i>
<i>9 Feb 12</i>	<i>Accuracy and Precision test</i> [For RPC or Simulator Solution Change Place Label Here] <i>Average .098 STD Dev 0.000</i> <i>Asum attempted Break report and received "Fatal error"</i>	<i>Tom DePant</i>
<i>3/24/12</i>	DataMaster Simulator Solution Lot # 12-38-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 11/17/2011 Expires: 11/17/2012 [For RPC]nge Place Label Here] OPENED: <i>3/24/12 LAFOUNTAIN</i> <i>SOLUTION CHANGE</i>	<i>LAFOUNTAIN</i>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
5/23/12	<u>DataMaster Simulator Solution</u> Lot # 12-64-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/17/2012 Expires: 02/17/2013 Place Label Here] OPENED: 5/23/12 LAFONTAIN SOLUTION CHANGE	LAFONTAIN CORPORAL
5/31/12	Reformed Software [For RPC or Simulator Solution Change Place Label Here] update and lock Change	A. B. G.
6/29/12	ROUTINE PERFORMANCE CHECK [For RPC or Simulator Solution Change Place Label Here] RPC	LAFONTAIN
10/31/12	ROUTINE PERFORMANCE CHECK [For RPC or Simulator Solution Change Place Label Here]	LAFONTAIN

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/16/12	[For RPC or Simulator Solution Change Place Label Here] APM Protocol performed	R Driscoll DMT TECH
2/26/12	[For R DataMaster Simulator Solution lace Label Here] Lot # 12-72-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/24/2012 Expires: 04/24/2013 OPENED: .	D. Pant
2/21/13	[For RPC or Simulator Solution Change Place Label Here] Touch Screen not working. Intermittent so tried a new ribbon cable with no luck. Found screen too tight so loosened mounts and now working to spec. RCO 2/1/13	R. Driscoll
2/21/13	DataMaster Simulator Solution Lot # 12-72-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/24/2012 Expires: 04/24/2013 OPENED: 2/21/13 LAFOUNTAIN SOLUTION CHANGE / RPC	LAFOUNTAIN

DATAMASTER DMT MAINTENANCE LOG

Milton PD
Department/Agency

100159
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10 Jun 13	<i>RPC PASSED</i> [For RPC or Simulator Solution Change Place Label Here]	<i>D. Pont</i>
14 Jun 13	DataMaster Simulator Solution Lot # 12-88-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 11/9/2012 Expires: 11/9/2013 [For RPC or Simulator Solution Change Place Label Here] OPENED: <i>RPC PASSED</i>	<i>D. Pont</i>
10/28/13	DataMaster Simulator Solution Lot # 13-03-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 04/03/2013 Expires: 04/03/2014 [For RPC or Simulator Solution Change Place Label Here] OPENED: <i>10/28/13 LA FOUNTAIN</i> <i>RPC PASSED</i>	<i>LA FOUNTAIN</i>
11/19/13	<i>Completed APM Protocol</i> [For RPC or Simulator Solution Change Place Label Here]	<i>R. Driscoll</i> <i>DMT TECH</i>

DATAMASTER DMT MAINTENANCE LOG

MILTON P.D.

Department/Agency

100159

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/24/14	DataMaster Simulator Solution [For R Lot # 13-03-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 04/03/2013 Expires: 04/03/2014 OPENED: 2/24/14 LAFOUNTAIN RPC PASSED	LAFOUNTAIN DMT SUPERVISOR
6/26/14	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 14-01-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 1/9/2014 Expires: 1/9/2015 OPENED: 6/26/14 LAFOUNTAIN RPC PASSED	LAFOUNTAIN
9/27/14	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 14-01-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 1/9/2014 Expires: 1/9/2015 OPENED: 9/27/14 LARENTE Simulator Solution Change	Larente DMT Supervisor
10/29/14	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 14-10-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 7/14/2014 Expires: 7/14/2015 OPENED: 10-29-14 Larente RPC PASSED	Larente DMT Supervisor

DATAMASTER DMT MAINTENANCE LOG

M.H. 20

Department/Agency

10059

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
12/8/14	Completed AQM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. D. Dussell DMT TECH
2/24/15	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 14-14-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/16/2014 Expires: 10/16/2015 OPENED: 2-24-15 Larente RPC PASSED	Larente DMT Supervisor
6/12/15	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 14-14-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/16/2014 Expires: 10/16/2015 OPENED: 6-12-15 LARENTE RPC PASSED	Larente DMT Supervisor
9/3/15	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 9/3/15 LAFOUNTAIN OPENED: SIMULATOR SOLUTION CHANGE	LAFOUNTAIN DMT SUPERVISOR

DATAMASTER DMT MAINTENANCE LOG

Milton PD

Department/Agency

100159

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/20/15	Downloaded data to USB drive, uploaded new [For RPC or Simulator Solution Change Place Label Here] software, and installed new I/O PCB. Not connected to the DPS server as there currently is no Ethernet port available	R. Driscoll DMT Tech.
10/28/15	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 15-12-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/28/2015 Expires: 07/28/2016 OPENED: 10-28-15 - Larente RPC PASSED	LARENTE DMT SUPERVISOR
12/10/15	Completed ARM protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT Tech.
12-31-15	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 15-12-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/28/2015 Expires: 07/28/2016 OPENED: 12-31-15 LARENTE	LARENTE DMT SUPERVISOR

DATAMASTER DMT MAINTENANCE LOG

MILTON P.D.

Department/Agency

100159

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2-9-16	[For RPC or Simulator Solution Change Place Label Here] DataMaster External Standard Solution Lot # 15-24-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 11/12/2015 Expires: 11/12/2016 OPENED: 2-9-16 Larente RPC PASSED	Larente DMT Supervisor
3-31-16	[For RPC or Simulator Solution Change Place Label Here] DataMaster External Standard Solution Lot # 15-25-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/30/2015 Expires: 11/30/2016 OPENED: 3-31-16 LARENTIE Simulator Solution Change	Larente DMT Supervisor
5/2/16	[For RPC or Simulator Solution Change Place Label Here] DataMaster External Standard Solution Lot # 15-24-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 11/12/2015 Expires: 11/12/2016 OPENED: 5/2/16 LAFOUNTAIN SOLUTION CHANGE	LAFOUNTAIN DMT SUPERVISOR
6/13/16	[For RPC or Simulator Solution Change Place Label Here] DataMaster External Standard Solution Lot # 15-25-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/30/2015 Expires: 11/30/2016 OPENED: 6/13/16 LAFOUNTAIN RPC PASSED	LAFOUNTAIN DMT SUPERVISOR

DATAMASTER DMT MAINTENANCE LOG

MILTON P.D.

Department/Agency

100159

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
8/3/16	[For RPC or Simulator Solution Change Place Label Here] DataMaster External Standard Solution Lot # 16-05-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 4/19/2016 Expires: 4/19/2017 OPENED: 8/3/16 LAFONTAIN SOLUTION CHANGE	LAFONTAIN DMT SUPERVISOR
10/3/16	[For RPC or Simulator Solution Change Place Label Here] DMT External Standard Solution Lot # 16-15-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 7/20/2016 Expires: 7/20/2017 OPENED: 10-3-16 LARENTE RPC PAISED	LARENTE DMT Supervisor
11-15-16	[For RPC or Simulator Solution Change Place Label Here] DMT External Standard Solution Lot # 16-15-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 7/20/2016 Expires: 7/20/2017 OPENED: 11-15-16 LARENTE SOLUTION CHANGE	LARENTE DMT SUPERVISOR
12-19-16	[For RPC or Simulator Solution Change Place Label Here] DMT External Standard Solution Lot # 16-23-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/19/2016 Expires: 10/19/2017 OPENED: 12-19-16 Larente Solution Change	Larente DMT Supervisor

DATAMASTER DMT MAINTENANCE LOG

MILTON P.D.

Department/Agency

100159

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
12/29/16	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. J. Russell DMT Tech.
01/05/17	DMT External Standard Solution Lot # 16-23-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/19/2016 Expires: 10/19/2017 OPENED: 01/05/17 LAFOUNTAIN [For RPC or Simulator Solution Change Place Label Here] SOLUTION CHANGE	LAFOUNTAIN DMT SUPERVISOR
01/20/17	[For RPC or Simulator Solution Change Place Label Here] DMT External Standard Solution Lot # 16-35-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/14/2016 Expires: 11/14/2017 OPENED: 1/20/17 LAFOUNTAIN SOLUTION CHANGE	LAFOUNTAIN DMT SUPERVISOR
1/24/17	[For RPC or Simulator Solution Change Place Label Here] COMPLETED ACCURACY + PRECISION CHECK	LAFOUNTAIN DMT SUPERVISOR

DATA MASTER DMT MAINTENANCE LOG

MILTON P.D.

Department/Agency

100159

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2-24-17	[For R DMT External Standard Solution Lot # 16-35-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/14/2016 Expires: 11/14/2017 OPENED: 2/24/17 Larente place Label Here] RPC Passed	Larente DMT Supervisor
3/16/17	[For RPC DMT Simulator Solution Lot # 17-10-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/06/2017 Expires: 02/06/2018 OPENED: 3/16/17 RCD e Label Here] Install post simulator update	R. Driscoll DMT Tech.
4/8/17	[For R DMT Simulator Solution Lot # 17-10-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/06/2017 Expires: 02/06/2018 OPENED: 4-8-17 Larente place Label Here] Simulator Solution Change	Larente DMT Supervisor
6/26/17	[For RPC DMT Simulator Solution Lot # 17-10-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/06/2017 Expires: 02/06/2018 OPENED: 6-26-17 LARENTE e Label Here] RPC PASSED	Larente DMT Supervisor

DATAMASTER DMT MAINTENANCE LOG

Milton ID

Department/Agency

100159

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/7/2017	DMT Simulator Solution Lot # 17-20-100 EtOHaq Certified Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 07/07/2017 Expires: 07/07/2018 [For RPC or Simulator Solution Change Place Label Here] OPENED: 10-7-2017 LARENTE RPC PASSED	LARENTE DMT Supervisor
10/27/17	Installed version 2.08 PIC chip [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFL
12/11/2017	DMT Simulator Solution Lot # 17-20-100 EtOHaq Certified Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 07/07/2017 Expires: 07/07/2018 [For RPC or Simulator Solution Change Place Label Here] OPENED: 12-11-17 LARENTE Simulator Solution Change	LARENTE DMT supervisor
2/11/18	DMT External Std Solution Lot: 17-24-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 11/14/17 Exp Date: 11/14/18 [For RPC or Simulator Solution Change Place Label Here] Opened: 2-11-18 Larente RPC Passed	Larente DMT Supervisor

DATAMASTER DMT MAINTENANCE LOG

Milton PD

Department/Agency

100159

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/23/18	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	R. Dineen VFL
6/3/18	DMT External Std Solution Lot: 17-24-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 11/14/17 Exp Date: 11/14/18 Opened: 6-3-18 LARENTE [For RPC or Simulator Solution Change Place Label Here] RPC PASSED	Larente DMT Supervisor
10/20/18	DMT External Std Solution Lot: 18-01-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 1/29/18 Exp Date: 01/29/19 Opened: 10/20/2018 LARENTE [For RPC or Simulator Solution Change Place Label Here] RPC Passed	Larente DMT Supervisor
1/11/19	DMT External Std Solution Lot: 18-01-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 1/29/18 Exp Date: 01/29/19 Opened: 1-11-19 LARENTE [For RPC or Simulator Solution Change Place Label Here] Simulator Solution Change	Larente DMT Supervisor

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/18/19	[For RPC or Simulator Solution Change Place Label Here] Accuracy + Precision TEST	LA FOUNTAIN DMT SUPERVISOR
1/25/19	Accuracy + Precision TEST	Locke DMT Supervisor
1/28/19	DMT External Std Solution Lot: 18-19-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 12/19/18 Exp Date: 12/19/19 : Label Here] Opened: 1/28/19 JONES	Locke JONES DMT Supervisors
2/19/19	Completed ARM Protocol [For RPC or Simulator Solution Change Place Label Here]	J Dulatte (JFL)
3/18/19	Accuracy + Precision Test	Locke
5/4/19	DMT External Std Solution Lot: 18-19-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 .abel Here] Prep Date: 12/19/18 Exp Date: 12/19/19 Opened: 5/4/19 Jones Simulator Solution change	Jones DMT Supervisor

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/21/19	DMT External Std Solution Lot: 19-06-100 EtOH Certified Conc. 0.098 Range: 0.093 - 0.103 Prep Date: 4/3/19 Exp Date: 4/3/20 Opened: 6/21/19 Jones [For RPC or Simulator Solution Change Place Label Here] RPC / Solution Change Passed	Jones DMT Supervisor
10/3/19	DMT External Std Solution Lot: 19-06-100 EtOH Certified Conc. 0.098 Range: 0.093 - 0.103 Prep Date: 4/3/19 Exp Date: 4/3/20 Opened: 10/3/19 Jones. [For RPC or Simulator Solution Change Place Label Here] RPC / Solution change	Jones DMT Supervisor
2/14/2020	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here] - 1st attempt had RF detected error during sim test - 2nd attempt passed	S. Stratton (VFL)
2/15/20	DMT External Std Solution Lot: 20-03-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 1/29/20 Exp Date: 01/29/21 Opened: 2/15/2020 NJ [For RPC or Simulator Solution Change Place Label Here] RPC Passed	N. Jones

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6-21-20	DMT External Std Solution [For RPC Lot: 20-03-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 1/29/20 Exp Date: 01/29/21 Opened: ce Label Here] RPC-Passw (P)	Sgt. P. Louie
6/22/2020	DMT supervisor reported touch screen [For RPC or Simulator Solution Change Place Label Here] was unresponsive. Obtained blank errors during check-in. Sim ruby replaced and connections secured. AEP run followed by successful check-in	J Dukette (VFL)
7/15/2020	Installed DMT software update [For RPC or Simulator Solution Change Place Label Here]	S Stratton (VFL)
8/31/2020	VFL staff responded to reports of errors. [For RPC or Simulator Solution Change Place Label Here] Could not reproduce any errors except printer error. DMT left in service 8/31/2020	J Dukette (VFL)

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
8/31/20	<i>Solution Change</i> [For RPC or Simulator Solution Change Place Label Here] DMT External Std Solution Lot: 20-10-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/25/20 Exp Date: 06/25/22 Opened: 8/31/20 <i>(P)</i>	P. Locke (SGT)
11-9-20	<i>RPC - Suck Back Error</i> <i>RPC #2 -> PASS</i> DMT External Std Solution Lot: 20-10-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/25/20 Exp Date: 06/25/22 Opened:	P. Locke SGT,
1/11/2021	<i>VFL staff responded to report of screen problems.</i> [For RPC or Simulator Solution Change Place Label Here] <i>DMT was taken to the VFL for repairs.</i>	S. Stratton (VFL)
1/26/2021	DMT External Std Solution [For RPC or Simulator Solution Change Place Label Here] Lot: 20-11-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/9/20 Exp Date: 07/09/22 Opened: 1/26/2021 <i>8018</i> <i>Performed installation protocol. DMT successfully reinstalled.</i>	S. Stratton (VFL)
2/12/2021	<i>Attempted to perform APM protocol but APM failed twice due to cal check errors.</i> [For RPC or Simulator Solution Change Place Label Here] <i>DMT taken back to the VFL for repairs.</i>	S. Stratton (VFL)

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/22/2021	DMT External Std Solution Lot: 20-15-100 EtOH Certified Conc. 0.100 [For RPC Range: 0.095 - 0.105 Label Here] Prep Date: 11/20/20 Exp Date: 11/20/22 Opened: 2/22/2021 8/18 Performed installation protocol. DMT successfully reinstalled. Completed APM protocol.	S. Stratton (VFL)
3/18/21	DMT External Std Solution Lot: 21-06-100 EtOH Certified Conc. 0.100 [For RPC Range: 0.095 - 0.105 e Label Here] Prep Date: 3/25/21 Exp Date: 03/25/23 Opened: 8/18 5/14/2021 Performed successful installation protocol.	J. Dukette (VFL)
10-12-21	DMT External Std Solution Lot: 21-04-100 EtOH Certified Conc. 0.100 [For RPC Range: 0.095 - 0.105 Place Label Here] Prep Date: 2/12/21 Exp Date: 02/12/23 Opened:	P. Locke

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
12/8/21	DMT External Std Solution [For RPC Lot: 21-12-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 10/8/21 Exp Date: 10/08/23 Opened: 12/8/21] Install post TSI	J. Dukette (VFL)
4/4/2022	DMT External Std Solution [For RPC Lot: 22-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/7/22 Exp Date: 03/07/24 Opened: 4/4/2022] Replaced the line conditioner. Installed post TSI	S. Stratton (VFL)
	DMT External Std Solution [For RPC Lot: 21-12-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 10/8/21 Exp Date: 10/08/23 Opened: 7/2/22 NS] RPC Passed	N. Jeng
10/9/22	DMT External Std Solution [For RPC Lot: 22-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/7/22 Exp Date: 03/07/24 Opened: 10-9-22 PL]	P. Locke
10/17/22	DMT External Std Solution [For RPC Lot: 21-12-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 10/8/21 Exp Date: 10/08/23 Opened: 10/17/22 @] solution change	P. Locke

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/8/23	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
	DMT External Std Solution Lot: 23-01-100 EtOH Certified Conc. 0.100 Rep Date: 1/25/23 Exp Date: 01/25/25 Opened: 3/2/23 NS RPC 3/2/23 Pass	A. Jones
7-1-23	DMT External Std Solution Lot: 23-01-100 EtOH Certified Conc. 0.100 Rep Date: 1/25/23 Exp Date: 01/25/25 Opened: 7-1-23 PL RPC PL	Locke, P
10-12-23	DMT External Std Solution Lot: 23-08-100 EtOH Certified Conc. 0.099 Prep Date: 7/19/23 Exp Date: 07/19/25 Opened: 10-12-23 PL RPC PL	Locke, P

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/6/24	Rebooted instrument to clear 'Busy' message. Completed APM [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
2/7/24	RPC DMT External Std Solution Lot: 23-08-100 EtOH Certified Conc. 0.099 Prep Date: 7/19/23 Exp Date: 07/19/25 Opened: 21 2/7/24	Locke @
6/14/24	RPC DMT Quality Control Solution Lot: 24-07-100 EtOH Certified Conc. 0.100 Prep Date: 5/29/24 Exp Date: 05/29/26 Opened: 6/14/24 @	P. Locke F
1/24/25	DMT removed for refurb [For RPC or Simulator Solution Change Place Label Here]	J Dukette (VFL)
	[For RPC or Simulator Solution Change Place Label Here]	

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	