

DATAMASTER DMT MAINTENANCE LOG

Rochester

100160

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/28/11	<p>DataMaster Simulator Solution Lot # 12-13-100 EtOH_{aq} Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 09/09/2011 Expires: 09/09/2012 [Range Place Label Here] OPENED: 10/28/11</p> <p style="text-align: center;">Install Per <i>SAW 10/28/11</i> Procedure</p>	Steven Harrods
11/08/11	<p>[For RPC or Simulator Solution Change Place Label Here] Updated Software corrected time</p>	Steven Harrods
2/8/12	<p>DataMaster Simulator Solution Lot # 12-13-100 EtOH_{aq} Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 09/09/2011 Expires: 09/09/2012 [For RPC Range Place Label Here] OPENED: 2/8/12 <i>THS</i></p>	Simpson Tom
5/17/12	<p>DataMaster Simulator Solution Lot # 12-64-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/17/2012 Expires: 02/17/2013 [For RPC Range Place Label Here] OPENED:</p> <p style="text-align: center;">Solution Change old one went out of range (.092) <i>(THS)</i></p>	Simpson Tom

DATAMASTER DMT MAINTENANCE LOG

Rochester Constable
Department/Agency

100160
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/20/12	<p>DataMaster Simulator Solution Lot # 12-70-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/17/2012 Expires: 04/17/2013 [For] Place Label Here]</p> <p>OPENED: 6/20/12 (TAS)</p> <p style="text-align: center;">RPC</p>	Tom Simpson DM Supv 1st Constable
8/28/12	<p style="color: blue;">APM Completed + Software install/update</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	R. Driscoll
10/21/12	<p>DataMaster Simulator Solution Lot # 12-58-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 02/09/2012 Expires: 02/09/2013 [For RPC] Label Here]</p> <p>OPENED: 10/21/12 (TAS)</p> <p style="text-align: center;">RPC</p>	Tom Simpson
1/22/13	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">Adj clock for EST</p>	Tom Simpson

DATAMASTER DMT MAINTENANCE LOG

Rochester Constables
Department/Agency

100160
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/15/13	<p>DataMaster Simulator Solution Lot # 12-70-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/17/2012 Expires: 04/17/2013</p> <p>[For RPC Label Here]</p> <p>OPENED: _____</p> <p style="font-size: 2em; text-align: center;">Test R.P.C.</p>	<p>TOM SIMPSON</p>
6/20/13	<p>DataMaster Simulator Solution Lot # 12-80-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/5/2012 Expires: 7/5/2013</p> <p>[For RF ace Label Here]</p> <p>OPENED: _____</p> <p style="font-size: 2em; text-align: center;">Test /RPC</p>	<p>TOM SIMPSON</p>
8/29/13	<p style="color: blue; text-align: center;">Completed APM Protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p style="color: blue;">R. Driscoll DMT TECH</p>
11/1/13	<p>DataMaster Simulator Solution Lot # 12-89-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 11/21/2012 Expires: 11/21/2013</p> <p>[For RPC Label Here]</p> <p>OPENED: 11/1/13</p> <p style="font-size: 2em; text-align: center;">TEST R.P.C.</p>	<p>TOM SIMPSON</p>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/8/13	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center; font-size: 1.2em;">Reset clock for EST</p>	TOM SIMPSON
2/25/14	<p>DataMaster Simulator Solution Lot # 14-01-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 1/9/2014 Expires: 1/9/2015</p> <p>[For _____ Place Label Here]</p> <p>OPENED: 2/25/14 TMS RPC</p>	TOM SIMPSON
6/18/14	<p>DataMaster Simulator Solution Lot # 14-01-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 1/9/2014 Expires: 1/9/2015</p> <p>[For _____ Place Label Here]</p> <p>OPENED: 6/18/14 TMS</p> <p style="text-align: center; font-size: 1.2em;">Adj. clock RPC</p>	TOM SIMPSON
9/17/14	<p>During APM protocol flow rate was 7.65 Lpm</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>Replaced ^{RO} 9/17/14 Installed adjustable flow rate valve, but do not have 1.5L ^{syringe} pump to set volume. Will come back to set volume. RO Set Volume to 1.5L with syringe RO 9/19/14 and completed APM protocol</p>	R. Driscoll DMT TECH

DATAMASTER DMT MAINTENANCE LOG

ROCHESTER CONSTABLES

Department/Agency

100160

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/19/14	<p>DataMaster Simulator Solution Lot # 14-05-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 6/19/2014 Expires: 6/19/2015</p> <p>OPENED:</p> <p>RPC #1 failed RF Detected x2 RPC #2 PASSED all tests (TAS)</p>	TOM SIMPSON
11/12	<p>Adj clock (EST OK but 7 min slow)</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	Tom SIMPSON
2/19/15	<p>DataMaster Simulator Solution Lot # 14-14-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/16/2014 Expires: 10/16/2015</p> <p>OPENED: 2/19/15 TAS</p> <p>RPC</p> <p>Adjust clock (5 min slow) (TAS)</p>	Tom SIMPSON
2/24/15	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>SAME AS ABOVE</p> <p>RPC error showed lot # 14-14-100</p> <p>SIM SOLUTION MADE TO CORRECT TO 14-14-100</p>	Tom SIMPSON

DATAMASTER DMT MAINTENANCE LOG

ROCH CONS

100160

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/10/15	Adj clock for DST [For RPC or Simulator Solution Change Place Label Here]	Tom Simpson
6/25/15	<p>DataMaster Simulator Solution Lot # 15-01-100 EtOHaq Conc: 0.098 g/210L Range: 0.093 - 0.103 g/210L Preparation Date: 1/21/2015 Expires: 1/21/2016</p> <p>[For RP] ce Label Here]</p> <p>OPENED: 6/25/15 TMS</p> <p style="text-align: center;">RPC</p>	Tom Simpson
10/25/15	<p>DataMaster Simulator Solution Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016</p> <p>[For RP] ace Label Here]</p> <p>OPENED: 10/25/15 TMS</p> <p style="text-align: center;">RPC</p>	Tom Simpson
11/3/15	<p>Completed ARM protocol. Downloaded data to [For RPC or Simulator Solution Change Place Label Here]</p> <p>USB, uploaded new software, installed new I/O PCB, and adjusted clock for day light Savings. No active Ethernet port available at this time.</p>	R. Driscoll DMT Tech.

DATAMASTER DMT MAINTENANCE LOG

ROCHESTER COMITABLES
Department/Agency

100160
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/14/16	<p>DataMaster Simulator Solution Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016 [For RF Place Label Here] OPENED: 02/14/16</p> <p style="text-align: center; font-size: 1.5em;">TEST / RPC</p>	<p style="font-size: 1.5em;">TSM SIMPSON</p>
6/26/16	<p>DataMaster External Standard Solution Lot # 15-23-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 10/29/2015 Expires: 10/29/2016 [For Place Label Here] OPENED: 6/26/16 TAS</p> <p style="text-align: center; font-size: 1.5em;">TEST / RPC</p>	<p style="font-size: 1.5em;">Tom SIMPSON</p>
6/28/16	<p style="text-align: center; font-size: 1.5em;">TEST / RPC</p>	<p style="font-size: 1.5em;">Mark Behrle</p>
10/25/16	<p>DMT External Standard Solution Lot # 16-21-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 9/13/2016 Expires: 9/13/2017 [Place Label Here] OPENED: 10/25/16 MAB</p> <p style="text-align: center; font-size: 1.5em;">TEST / RPC</p>	<p style="font-size: 1.5em;">Mark Behrle</p>
10/25/16	<p style="text-align: center; font-size: 1.5em;">TEST / RPC</p>	<p style="font-size: 1.5em;">Mark Behrle</p>
11/3/16	<p style="text-align: center; color: blue; font-size: 1.2em;">Completed ARM Protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p style="color: blue; font-size: 1.2em;">R. Driscoll DMT Tech</p>
2/27/17	<p style="text-align: center; font-size: 1.5em;">TEST RPC</p>	<p style="font-size: 1.5em;">Mark Behrle</p>

DATAMASTER DMT MAINTENANCE LOG

Rochester Constables
Department/Agency

100160
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/27/17	<p>DMT Simulator Solution Lot # 17-13-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/28/2017 Expires: 02/28/2018 OPENED: 3/27/17</p> <p style="text-align: center;">INSTALLATION AFTER SIM UPGRADE</p>	D. JARLAK VFL
1/1/23	<p style="text-align: center;">TEST RPC</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	Mark Belisle
11/29/17	<p style="text-align: center;">Replaced PIC chip w/ version 2.08</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	R. D. Sc. II VFL
2/11/18	<p style="text-align: center;">DMT External Std Solution Lot: 17-24-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 11/14/17 Exp Date: 11/14/18 Opened:</p> <p>[For _____ Place Label Here]</p> <p style="text-align: center;">RPC</p>	Mark Belisle
2/11/18		

DATAMASTER DMT MAINTENANCE LOG

Rochester Constable
Department/Agency

100160
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/21/18	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFC
6/20/18	RPC	M. Belisle
6/13/18	RPC [For RPC or Simulator Solution Change Place Label Here]	out of range
10/13/18	see below	
10/13/18	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RPC] Lot: 18-01-100 EtOH Certified Conc. 0.101 [Label Here]</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 1/29/18 Exp Date: 01/29/19</p> <p>Opened: <u>10/13/18</u></p>	
2/8/19	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RPC] Lot: 18-19-100 EtOH Certified Conc. 0.101 [Label Here]</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 12/19/18 Exp Date: 12/19/19</p> <p>Opened:</p>	

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/15/19	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	J Dukette (UFC)
4/30/19	DMT removed from agency [For RPC or Simulator Solution Change Place Label Here]	J Dukette (UFC)
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	