

**DATAMASTER DMT OPERATOR USE LOG**

Rochester

100160

Department/Agency

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
10/28/11	Steven Harrois	Install	.100	
10/29/11	Collier, A (TAS)	Matt Manning	.102 .097 →	
2/8/12	SIMPSON, T	Test RPC	.099	
5/12/12	SIMPSON, T	Solution change		
6/20/12	SIMPSON, T	TEST RPC	.099	
8/28/12	R. Driscoll	TEST / ARM	.095	—
10/21/12	SIMPSON, Tom	RPC	.099	—
1/19/13	Giolito, Robert	Chester Baxter	.095	
1/19/13	Giolito, Robert	Chester Baxter		
2/15/13	SIMPSON, T	RPC	.098	
2/21/13	CHECK IN TEST Tom PER R.D. SIMPSON →		.097	
3/2/13	CHECK IN TEST Per A.B. w/all 4 breath types →		.097	
3/8/13	Adj clock for D.S.T.	(TAS) m741	—	
6/20/13	SIMPSON, TOM	RPC		
8/29/13	R. Driscoll	TEST / ARM	.097	—
11/1/13	SIMPSON, T	RPC	.096	—
11/8	"	reset clock for EST	—	—
2/24/14	"	TEST / RPC	.098	
"	reset clock correct 5 min off			
6-7-14	BROUILLARD	STEPHEN MACKAY	.094	

# DATAMASTER DMT OPERATOR USE LOG

ROCB CONS

Department/Agency

100160

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
4/18/14	Tom SIMPSON	TEST RPC	.098	—
9/17/14	R. Driscoll	TEST / AOM	.096	—
10/19/14	Tom Simpson	TEST RPC	.098	—
10/19/14	" "	TEST RPC	.098	—
2/19/15	Tom Simpson <sup>reset</sup> <sub>clock</sub>	TEST RPE	.098	—
2/24/15	Tom Simpson	Correct lot # via SIM SOL change	.097	←
3/10/15	Tom Simpson	Adj clock (DST)		
6/25/15	Tom Simpson	Test RPC	.098	
8.14.15	N. Allwood	RICHARD MACLAY	.098	
10/25/15	Tom Simpson	RPC Test	.099	
11/3/15	R. Driscoll	TEST / AOM	.098	—
1.15.16	N. Allwood	Andrew Buro	.099	
2/14/16	TOM SIMPSON	TEST / RPC	.099	
4/27/16	Mark Belisle	TEST RUSTIC		
8-11-16	Helfant	Richard Gibson	.098	
9-11-16	m BUSIER	Calder Wilmot	.098	
10/25/16	Mark Belisle	RPC Test	.098	
11/3/16	R. Driscoll	TEST / AOM	.098	—
02/03/17	Jonathan Duncan	Brad Champagne Test	.098	
2/20/17	M. BUSIER	MATT E. MANNING	.098	

**DATAMASTER DMT OPERATOR USE LOG**

Roch Co. V.S

Department/Agency

160160

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
2/20	M. BUSIER	TYLER AINSWORTH	.098	
2/27	m Belisle	test RPC	.098	
3/18	M. BUSIER	KEVIN M. DOWDON	.098	
6/11	m Belisle	TEST RPC	.099	
9/1	M. BUSIER	JOAN THRESHER	.098	
9/9	M. BUSIER	HENRY THREATCH	.097	
10/23	m Belisle	<del>TEST</del> RPC	.097	
2/11/18	m Belisle	Solution change RPC	.100	
2/18/18	J Pregent	Mark Hammond	.100	
2/21/18	R. Daisell	TEST / ARM	0.099	—
6/20/18	m Belisle	test RPC	0.097	
07/03/18	J Duncan	Allison Brown	.100	
10/13/18	m Belisle	test RPC	.094	.094
10/13/18	m Belisle	Solution change test RPC	.099	
12/18/18	M. BUSIER	Robert W. Krattli	.098	
2/8/19	m Belisle	Solution change test RPC	.100	
2/15/19	J Dukette	Test / ARM	0.099	