

# ANNUAL PREVENTATIVE MAINTENANCE



DataMaster DMT: 100160  
Location: ST. ALBANS PD  
Calibration Date: 05/12/2020  
Certification Date: 05/12/2020  
Installation Date: 06/22/2020  
APM Date: 06/22/2020  
Performed by: JEFFERY DUKETTE

## Diagnostic Results

### VERSIONS

DMT: 2.05  
PIC: 2.08  
Modem: 2.6  
Questions: 2.1

### TEMPERATURES

Sample Chamber = 48.8°C  
Breath Tube = 48.1°C  
Digital Sim = 34.0°C

### SETTINGS

Lamp Voltage = 1.64 V  
Cooler Voltage = 1.66 V  
Bias Voltage = 80 V  
Chopper Freq = 545 Hz

### PUMP INFO

Flow Rate = 4.830 L/M

### DETECTOR INFO

PUMP	ON	OFF
MAX(V)	0.0624	0.0673
MIN(V)	0.0606	0.0642

### FILTER INFO

Filter 1	0.066	Zero = true
Filter 2	0.370	Zero = true
Filter 3	1.565	Zero = true

### CALIBRATION CHECK

Xq = 0.087 0.69%

**APM PASSED**

## Simulator Vapor Test


Concentration = 0.101 g/210L  
Lot # = 20-05-100  
Range = 0.096 - 0.106  
Simulator Vapor = 0.100 g/210L

## RF Detection Test

Passed

## Sample Acceptance Test

Passed



Performed by

*[Signature]*

Date

06/22/2020

Reviewed by

*[Signature]*

Date

6/25/2020

Simulator Temperature Check

Simulator Serial Number:

MP3505

DMT Serial Number / Simulator Location:

100160 / St Albans PD

NIST Traceable Thermometer Serial Number:

192420790

As Found

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator	34.08	34.01	0.07	Y

\* 34.0 +/- 0.2 degree Celsius acceptable range

Was temperature adjustment performed?

Y / N

As Left (if applicable)

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator				

Notes:

Check Performed By:

[Signature]

Date:

6/22/2020

Reviewed by:

[Signature]

Date:

6/25/2020