

INSTALLATION REPORT

DataMaster DMT: 100160
Location: ST. ALBANS PD
Calibration Date: 12/24/2019
Certification Date: 12/24/2019
Installation Date: 01/03/2020
Installed by: JEFFERY DUKETTE



Default Options Loaded

Diagnostic Results

VERSIONS
DMT: 2.05
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.7°C
Breath Tube = 48.1°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.61 V
Cooler Voltage = 1.61 V
Bias Voltage = 80 V
Chopper Freq = 542 Hz

PUMP INFO

Flow Rate = 4.865 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX(V)	-0.1317	-0.1262
MIN(V)	-0.1337	-0.1272

FILTER INFO

Filter 1	-0.126	Zero = true
Filter 2	0.197	Zero = true
Filter 3	1.414	Zero = true

CALIBRATION CHECK

Xq = 0.113 0.55%

INSTALLATION PASSED

Accuracy and Precision Check


Concentration	= 0.101 g/210L
Lot #	= 19-16-100
Range	= 0.096 - 0.106
Average	= 0.100 g/210L
Std Dev	= 0.0005

RF Detection Test

Passed

Sample Acceptance Test

Passed



Performed by

[Signature]

Date 01/03/2020

Reviewed by

[Signature]

Date 1/6/2020

DMT Serial Number #100160

Page 1 of 1

01/03/2020 10:43 AM

100160

[Signature]

page 1 of 2

VERMONT FORENSIC LABORATORY

DMT Site Evaluation Checklist

Agency or Site Information

Name: St Albans PD

Address: 30 Lower Weldon St
St Albans, VT 05478

Phone: 802-524-2166

Fax: _____

Personnel Information

Head of Site: W. Couture

Current DMT Supervisor (s):
Michael Malinowski

Proposed DMT Supervisor(s):
N/A

DMT Environment

Limited Controlled Access:	<input checked="" type="radio"/> Y	<input type="radio"/> N
Low Activity Area:	<input type="radio"/> Y	<input checked="" type="radio"/> N **
Unlimited Access by Other Law Enforcement Personnel	<input checked="" type="radio"/> Y	<input type="radio"/> N
Access by VFL Technical Services	<input checked="" type="radio"/> Y	<input type="radio"/> N
Air Conditioned	<input checked="" type="radio"/> Y	<input type="radio"/> N
Room Temperature 65° - 85° F	<input checked="" type="radio"/> Y	<input type="radio"/> N
Adequate Ventilation	<input checked="" type="radio"/> Y	<input type="radio"/> N
Free of Cleaning Supplies	<input checked="" type="radio"/> Y	<input type="radio"/> N
Clean	<input checked="" type="radio"/> Y	<input type="radio"/> N

DMT Connection(s)

Power:
Min 15 Amp Y N

Limited low power sharing: Y N

Ethernet Available: Y N


Distance to Ethernet: 5' JSO 11/3/2020

DMT Site Maintenance Agreement

Will notify VFL Toxicology Section staff of maintenance and supply needs promptly: Y N

Will notify VFL Toxicology Section staff of changes in building conditions or moves promptly: Y N

Will notify VFL Toxicology Section staff of relevant personnel changes at site promptly: Y N

VFL Representative:	<u>Jeff Dukette</u> Print Name	 Signature	<u>11/2/2020</u> Date
Agency Representative:	<u>MARY</u> Print Name	 Signature	<u>1/2/22</u> Date

*Currently have a set up for a DMT. New power was run for DMT. Renewing our site evaluation agreement.

TOX_F200_4_v4_09102018
Approved by Lab Director

JSO 11/25/2019

** DMT is in secure location but is adjacent to an active egress used by officers. JSO 11/2/2020

100160

JSO

2