

# SOLUTION CHANGE

DataMaster DMT: 100160  
Location: ST. ALBANS PD  
Date: 04/13/2021  
Performed by: JOSEPH E THOMAS



## Accuracy and Precision Check

Concentration = 0.100 g/210L  
Lot # = 20-14-100  
Range = 0.095 - 0.105  
Average = 0.099 g/210L  
Std Dev = 0.0000

*SDS 4/14/2021*

Simulator Temperature: 34.0°C

Performed by

  
\_\_\_\_\_

Date

04/13/2021