

# SOLUTION CHANGE

DataMaster DMT: 100160  
Location: ST. ALBANS PD  
Date: 09/09/2021  
Performed by: JOSEPH E THOMAS



## Accuracy and Precision Check

Concentration = 0.100 g/210L  
Lot # = 21-11-100  
Range = 0.095 - 0.105  
Average = 0.098 g/210L  
Std Dev = 0.0004

*JSD*  
*9/9/21*

Simulator Temperature: 34.0°C

Performed by

  
\_\_\_\_\_

Date

09/09/2021