



Simulator Temperature Check

Simulator Serial Number: MP3505  
 DMT Serial Number / Simulator Location: 100160 / ST Albans PD  
 NIST Traceable Thermometer Serial Number: 192420790

As Found

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator	34.15	34.02	0.13	N

\* ± 0.10°C agreement between thermometer and simulator  
 \* 34.0 ± 0.2°C acceptable range

Was temperature adjustment performed?  Y / N

As Left (if applicable)

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator	33.98	34.00	0.02	Y

Notes:

Check Performed By:  Date: 1/25/22  
 Reviewed by:  Date: 1/28/22