

SOLUTION CHANGE

DataMaster DMT: 100160
Location: ST. ALBANS PD
Date: 01/08/2022
Performed by: JOSEPH E THOMAS



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 21-12-100
Range = 0.094 - 0.104
Average = 0.098 g/210L
Std Dev = 0.0007

JD
1/10/22

Simulator Temperature: 34.0°C

Performed by

A handwritten signature in black ink, appearing to read "Joseph E. Thomas", written over a horizontal line.

Date

01/08/2022