

ANNUAL PREVENTATIVE MAINTENANCE



DataMaster DMT: 100160
Location: ST. ALBANS PD
Calibration Date: 10/12/2022
Certification Date: 10/12/2022
Installation Date: 10/28/2022
APM Date: 01/19/2023
Performed by: JEFFERY DUKETTE

Diagnostic Results

VERSIONS
DMT: 2.06
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.6°C
Breath Tube = 48.1°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.67 V
Cooler Voltage = 1.66 V
Bias Voltage = 80 V
Chopper Freq = 543 Hz

PUMP INFO

Flow Rate = 5.088 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX (V)	-0.1236	-0.1179
MIN (V)	-0.1252	-0.1193

FILTER INFO

Filter 1	-0.117	Zero = true
Filter 2	0.196	Zero = true
Filter 3	1.397	Zero = true

CALIBRATION CHECK

Xq = 0.085 0.91%

APM PASSED

Simulator Vapor Test


Concentration = 0.098 g/210L
Lot # = 22-14-100
Range = 0.093 - 0.103
Simulator Vapor = 0.097 g/210L

RF Detection Test

Passed
RF detected

Sample Acceptance Test

Passed



Performed by



Date

01/19/2023

Reviewed by



Date

2/9/23

VERMONT FORENSIC LABORATORY

Simulator Temperature Check

Simulator Serial Number: MP3505
 DMT Serial Number / Simulator Location: 100160 / St Albans PD
 NIST Traceable Thermometer Serial Number: 21FMM02249

As Found

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator	33.84	34.01	0.17	N


* $\pm 0.10^{\circ}\text{C}$ agreement between thermometer and simulator
 * $34.0 \pm 0.2^{\circ}\text{C}$ acceptable range

Was temperature adjustment performed? Y / N

As Left (if applicable)

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator	33.99	34.00	0.01	Y

Notes:

Check Performed By:  Date: 11/19/23
 Reviewed by:  Date: 2/19/23