

# SOLUTION CHANGE

DataMaster DMT: 100160  
Location: ST. ALBANS PD  
Date: 01/13/2023  
Performed by: ANDREW G UNDERWOOD



## Accuracy and Precision Check

Concentration = 0.098 g/210L  
Lot # = 22-14-100  
Range = 0.093 - 0.103  
Average = 0.094 g/210L  
Std Dev = 0.0005

*SOS 1/13/2023*

Simulator Temperature: 34.0°C

Performed by

*Andrew Underwood*

Date

01/13/2023