

SOLUTION CHANGE

DataMaster DMT: 100160
Location: ST. ALBANS PD
Date: 05/20/2023
Performed by: KAYLIE M CADORETTE



Accuracy and Precision Check

Concentration = 0.100 g/210L
Lot # = 22-15-100
Range = 0.095 - 0.105
Average = 0.097 g/210L
Std Dev = 0.0007

JS
5/22/23

Simulator Temperature: 34.0°C

Performed by

Date

05/20/2023