

SOLUTION CHANGE

DataMaster DMT: 100160
Location: ST. ALBANS PD
Date: 08/05/2023
Performed by: JOSEPH E THOMAS



Accuracy and Precision Check

Concentration = 0.100 g/210L
Lot # = 22-15-100
Range = 0.095 - 0.105
Average = 0.098 g/210L
Std Dev = 0.0005

JSD 8/7/23

Simulator Temperature: 34.0°C

Performed by

A handwritten signature in blue ink, appearing to read 'J. E. Thomas', written over a horizontal line.

Date

08/05/2023

SOLUTION CHANGE

DataMaster DMT: 100160
Location: ST. ALBANS PD
Date: 08/05/2023
Performed by: JOSEPH E THOMAS



Accuracy and Precision Check

Concentration
Lot #
Range
Average
Std Dev
Interference Detected

Simulator Temperature:

Solution Change Failed

Performed by

Date

08/05/2023