

SOLUTION CHANGE

DataMaster DMT: 100160
Location: BERLIN PD
Date: 3/24/2025
Performed by: LEVI S WILLEY



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 24-24-100
Range = 0.094 - 0.104
Average = 0.100 g/210L
Std Dev = 0.0005

JSW
3/25/25

Simulator Temperature: 34.0°C

Performed by

MSW

Date

3/24/2025